

LB113526


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38435



A= 0.0000e+000

B= 2.5744e-004

C= 2.3390e-002

Rho= 0.9997943

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	90D
0.0	0.000	0.028	0.028	18	0.000	18					-
0.2	0.200	0.210	0.010	723	0.0 %	723					5
2.5	2.500	2.405	-0.095	9252	0.0 %	9252					-4
5.0	5.000	5.009	0.009	19367	0.0 %	19367					0
7.5	7.500	7.630	0.130	29548	0.0 %	29548					2
10.0	10.000	9.918	-0.082	38435	0.0 %	38435					-1

LB113526 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Mar 2021 11:26:40

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	18	-	0.0000	7470A	PPB	16 Mar 2021 11:28:35	S	Std
0.2 - 1	50.2	723	-	0.2000	7470A	PPB	16 Mar 2021 11:35:27	S	Std
2.5 - 1	52.5	9252	-	2.5000	7470A	PPB	16 Mar 2021 11:37:44	S	Std
5.0 - 1	55	19367	-	5.0000	7470A	PPB	16 Mar 2021 11:40:01	S	Std
7.5 - 1	57.5	29548	-	7.5000	7470A	PPB	16 Mar 2021 11:42:17	S	Std
10.0 - 1	510	38435	-	10.0000	7470A	PPB	16 Mar 2021 11:47:05	S	Std
ICV40 - 1	ICV40	14756	3.8221	-7470A	PPB	PPB	16 Mar 2021 11:49:48	U	SMPL
ICB40 - 1	ICB40	-31	0.0154	-7470A	PPB	PPB	16 Mar 2021 11:52:02	U	SMPL
CCV29 - 1	CCV29	20003	5.1729	-7470A	PPB	PPB	16 Mar 2021 11:54:17	U	SMPL
CCB29 - 1	CCB29	-38	0.0136	-7470A	PPB	PPB	16 Mar 2021 11:56:32	U	SMPL
CRA - 1	CRA	693	0.2018	-7470A	PPB	PPB	16 Mar 2021 11:58:47	U	SMPL
HighStd - 1	HighStd	38466	9.9260	-7470A	PPB	PPB	16 Mar 2021 12:01:01	U	SMPL
ChkStd - 1	ChkStd	27473	7.0959	-7470A	PPB	PPB	16 Mar 2021 12:05:54	U	SMPL
PB135168BL - 1	PBW	-57	0.0087	-7470A	PPB	PPB	16 Mar 2021 12:08:09	U	SMPL
PB135168BS - 1	LCSW	17140	4.4359	-7470A	PPB	PPB	16 Mar 2021 12:10:25	U	SMPL
M1666-01 - 1	1875-1	220	0.0800	-7470A	PPB	PPB	16 Mar 2021 12:12:41	U	SMPL
M1678-06 - 1	D2470-D2800(SOIL)-A	-37	0.0139	-7470A	PPB	PPB	16 Mar 2021 12:14:58	U	SMPL
M1678-06DUP - 1	D2470-D2800(SOIL)-ADUP	29	0.0309	-7470A	PPB	PPB	16 Mar 2021 12:17:14	U	SMPL
M1678-06MS - 1	D2470-D2800(SOIL)-AMS	14938	3.8690	-7470A	PPB	PPB	16 Mar 2021 12:19:31	U	SMPL
M1678-06MSD - 1	D2470-D2800(SOIL)-AMSD	14473	3.7493	-7470A	PPB	PPB	16 Mar 2021 12:21:46	U	SMPL
CCV30 - 1	CCV30	19323	4.9978	-7470A	PPB	PPB	16 Mar 2021 12:24:01	U	SMPL
CCB30 - 1	CCB30	-37	0.0139	-7470A	PPB	PPB	16 Mar 2021 12:26:16	U	SMPL
PB135169BL - 1	PBW	30	0.0311	-7470A	PPB	PPB	16 Mar 2021 12:28:31	U	SMPL
PB135169BS - 1	LCSW	14881	3.8543	-7470A	PPB	PPB	16 Mar 2021 12:30:46	U	SMPL
M1666-02 - 1	1858	-10	0.0208	-7470A	PPB	PPB	16 Mar 2021 12:33:01	U	SMPL
M1666-03 - 1	1875	109	0.0515	-7470A	PPB	PPB	16 Mar 2021 12:35:17	U	SMPL
M1666-04 - 1	1886	58	0.0383	-7470A	PPB	PPB	16 Mar 2021 12:37:32	U	SMPL
M1666-05 - 1	COMP-89-90	10	0.0260	-7470A	PPB	PPB	16 Mar 2021 12:39:48	U	SMPL
M1667-01 - 1	DRUM-TOTE-COMP	12	0.0265	-7470A	PPB	PPB	16 Mar 2021 12:42:04	U	SMPL
M1667-01DUP - 1	DRUM-TOTE-COMPDUP	24	0.0296	-7470A	PPB	PPB	16 Mar 2021 12:44:21	U	SMPL
M1667-01MS - 1	DRUM-TOTE-COMPMS	14642	3.7928	-7470A	PPB	PPB	16 Mar 2021 12:46:38	U	SMPL
M1667-01MSD - 1	DRUM-TOTE-COMPMSD	15278	3.9565	-7470A	PPB	PPB	16 Mar 2021 12:48:54	U	SMPL
CCV31 - 1	CCV31	19422	5.0233	-7470A	PPB	PPB	16 Mar 2021 12:51:09	U	SMPL
CCB31 - 1	CCB31	-36	0.0141	-7470A	PPB	PPB	16 Mar 2021 12:53:25	U	SMPL
PB135168TB - 1	PB135150TB	29	0.0309	-7470A	PPB	PPB	16 Mar 2021 12:55:40	U	SMPL
PB135169TB - 1	PB135151TB	20	0.0285	-7470A	PPB	PPB	16 Mar 2021 12:57:55	U	SMPL
M1678-06LX5 - 1		18	0.0280	-7470A	PPB	PPB	16 Mar 2021 13:00:11	U	SMPL
M1678-06A - 1		15277	3.9563	-7470A	PPB	PPB	16 Mar 2021 13:02:27	U	SMPL
M1667-01LX5 - 1		-11	0.0206	-7470A	PPB	PPB	16 Mar 2021 13:04:42	U	SMPL
M1667-01A - 1		14669	3.7997	-7470A	PPB	PPB	16 Mar 2021 13:06:58	U	SMPL
CCV32 - 1	CCV32	19221	4.9716	-7470A	PPB	PPB	16 Mar 2021 13:09:14	U	SMPL
CCB32 - 1	CCB32	-19	0.0185	-7470A	PPB	PPB	16 Mar 2021 13:11:31	U	SMPL