



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113528

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.9	03/15/2021	09:50
2	M1665-02		1	10	100.00	>212.0	>212.0	03/15/2021	11:10
3	M1665-02DUP		1	10	100.00	>212.0	>212.0	03/15/2021	11:40
4	M1665-03		1	11	100.00	>212.0	>212.0	03/15/2021	12:05
5	M1666-03		1	11	100.00	>212.0	>212.0	03/15/2021	12:35
6	M1666-04		1	12	100.00	>212.0	>212.0	03/15/2021	13:10
7	M1677-01		1	12	100.00	>212.0	>212.0	03/15/2021	13:40
8	M1677-03		1	11	100.00	>212.0	>212.0	03/15/2021	14:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp31521 WorkList ID : 147144 Department : Wet-Chemistry Date : 03-15-2021 09:17:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/19/2021	Water	M1665-02	Flash Point	Cool 4 deg C	PSEG03	H52	41739	03/12/2021	1010A
03/19/2021	Water	M1665-03	Flash Point	Cool 4 deg C	PSEG03	H52	30821	03/12/2021	1010A
03/19/2021	Water	M1666-03	Flash Point	Cool 4 deg C	PSEG03	H52	1875	03/12/2021	1010A
03/19/2021	Water	M1666-04	Flash Point	Cool 4 deg C	PSEG03	H52	1886	03/12/2021	1010A
03/19/2021	Water	M1677-01	Flash Point	Cool 4 deg C	TULL02	J42	D2470-D2800(WATER)	03/12/2021	1010A
03/19/2021	Water	M1677-03	Flash Point	Cool 4 deg C	TULL02		D2470-D2800(WATER)	03/12/2021	1010A

Date/Time 03/15/21 09:25
 Received by: KS (Wetchem)
 Relinquished by: RJ (EXT-196)

Date/Time 03/15/21 14:15
 Received by: RJ (EXT-196)
 Relinquished by: KS (Wetchem)