



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113529

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.9	03/15/2021	09:50
2	M1666-01		1	9	100.00	>212.0	>212.0	03/15/2021	10:15
3	M1666-01DUP		1	9	100.00	>212.0	>212.0	03/15/2021	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 113529

WorkList Name : fp31521-1 WorkList ID : 147146 Department : Wet-Chemistry Date : 03-15-2021 09:23:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/19/2021	Solid	M1666-01	Flash Point	Cool 4 deg C	PSEG03	H52	1875-1	03/12/2021	1010A

Date/Time 03/15/21 09:25
 Received by: KS (wetchen)
 Relinquished by: RJ (EFT-106b)

Date/Time 03/15/21 10:50
 Received by: RJ (EFT-106b)
 Relinquished by: KS (wetchen)