

LB113533


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41224



A= 0.0000e+000

B= 2.4207e-004 slope

C= 8.7335e-003 $\rightarrow$ intercept

Rho= 0.9998486

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.013	0.013	16	0.000	16					16
0.05	0.050	0.058	0.008	204	0.0 %	204					-7
0.20	0.200	0.187	-0.013	736	0.0 %	736					-3
2.50	2.500	2.419	-0.081	9955	0.0 %	9955					3
5.00	5.000	5.139	0.139	21192	0.0 %	21192					-1
7.50	7.500	7.447	-0.053	30728	0.0 %	30728					0
10.0	10.000	9.988	-0.012	41224	0.0 %	41224					

LB113533

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.05	16	-		0 ISM02.4	PPB	3/16/2021 15:33	Std
	0.05 S0.05	204	-		0.05 ISM02.4	PPB	3/16/2021 15:35	Std
	0.2 S0.2	736	-		0.2 ISM02.4	PPB	3/16/2021 15:37	Std
	2.5 S2.5	9955	-		2.5 ISM02.4	PPB	3/16/2021 15:40	Std
	5 S5.0	21192	-		5 ISM02.4	PPB	3/16/2021 15:42	Std
	7.5 S7.5	30728	-		7.5 ISM02.4	PPB	3/16/2021 15:49	Std
	10 S10.0	41224	-		10 ISM02.4	PPB	3/16/2021 15:59	Std
ICV	ICV	15509	3.763 -		ISM02.4	PPB	3/16/2021 16:02	SMPL
ICB	ICB	-21	0.0037 -		ISM02.4	PPB	3/16/2021 16:04	SMPL
CCV033	CCV033	20719	5.0242 -		ISM02.4	PPB	3/16/2021 16:06	SMPL
CCB033	CCB033	-31	0.0012 -		ISM02.4	PPB	3/16/2021 16:09	SMPL
PB135181BL	PBW181	16	0.0126 -		ISM02.4	PPB	3/16/2021 16:11	SMPL
M1657-01	MEW105	11	0.0114 -		ISM02.4	PPB	3/16/2021 16:13	SMPL
M1657-02	MEW107	17	0.0128 -		ISM02.4	PPB	3/16/2021 16:15	SMPL
M1657-03	MEW109	36	0.0174 -		ISM02.4	PPB	3/16/2021 16:18	SMPL
M1657-04	MEW109D	14	0.0121 -		ISM02.4	PPB	3/16/2021 16:20	SMPL
M1657-05	MEW109S	4217	1.0296 -		ISM02.4	PPB	3/16/2021 16:22	SMPL
CCV034	CCV034	21005	5.0935 -		ISM02.4	PPB	3/16/2021 16:24	SMPL
CCB034	CCB034	-21	0.0037 -		ISM02.4	PPB	3/16/2021 16:27	SMPL