



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113534

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/16/2021	14:05
2	CAL2	1	Water	NA	NA	20.1	7.00	03/16/2021	14:06
3	CAL3	1	Water	NA	NA	20.1	10.02	03/16/2021	14:07
4	ICV	1	Water	NA	NA	20.2	7.01	03/16/2021	14:08
5	CCV1	1	Water	NA	NA	20.2	2.01	03/16/2021	14:10
6	M1687-03	1	Solid	20.03	20	20.3	7.24	03/16/2021	14:11
7	M1687-03DUP	1	Solid	20.04	20	20.4	7.25	03/16/2021	14:12
8	M1687-04	1	Solid	20.03	20	20.4	6.68	03/16/2021	14:15
9	CCV2	1	Water	NA	NA	20.1	12.02	03/16/2021	14:16

2B 113534

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M1687 S WorkList ID : 147179 Department : Wet-Chemistry Date : 03-16-2021 13:57:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/15/2021	Solid	M1687-03	pH	Cool 4 deg C	PSEG03	H61	AUD-1052	03/15/2021	9045D
03/15/2021	Solid	M1687-04	pH	Cool 4 deg C	PSEG03	H61	AUD-1053	03/15/2021	9045D

Date/Time 03/16/21 14:00
Received by: JB (leg) (nm)
Relinquished by: JB (leg) (nm)

Date/Time 03/16/21 16:20
Received by: JB (leg) (nm)
Relinquished by: JB (leg) (nm)