



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB113536

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP90005
Turbidity Calibration std, 0NTU	WP90004
Turbidity Calibration std, 20NTU	WP90000
Turbidity Calibration std, 40NTU	WP89999
Turbidity Calibration std, 1NTU	WP90003
Turbidity Calibration std, 5NTU	WP90002
Turbidity Calibration - CCV std, 10 NTU	WP90001
10 NTU Standard 500 ml	W2752

Intercept: 0.3538

Slope: 0.9632

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.256	-0.10		03/16/2021	15:30
2	CAL2	1	1	1.250	0.93	7	03/16/2021	15:32
3	CAL3	5	1	5.360	5.20	-3.9	03/16/2021	15:35
4	CAL4	10	1	9.860	9.87	1.3	03/16/2021	15:37
5	CAL5	20	1	19.800	20.19	-0.9	03/16/2021	15:39
6	CAL6	40	1	38.800	39.92	0.2	03/16/2021	15:42



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB113536

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.100	10.119	03/16/2021	15:45
2	ICB		1	0.296	-0.060	03/16/2021	15:47
3	CCV1	10	1	9.940	9.952	03/16/2021	15:49
4	CCB1		1	0.280	-0.077	03/16/2021	15:52
5	LB113536BL		1	0.301	-0.055	03/16/2021	15:54
6	M1458-12		1	0.735	0.396	03/16/2021	15:57
7	M1561-36		1	16.200	16.452	03/16/2021	15:59
8	CCV2	10	1	9.770	9.776	03/16/2021	16:02
9	CCB2		1	0.286	-0.070	03/16/2021	16:04