



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB113538

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/16/2021	17:05
2	CAL2	1	Water	NA	NA	20.1	7.00	03/16/2021	17:06
3	CAL3	1	Water	NA	NA	20.1	10.03	03/16/2021	17:07
4	ICV	1	Water	NA	NA	20.2	6.98	03/16/2021	17:10
5	CCV1	1	Water	NA	NA	20.2	2.01	03/16/2021	17:11
6	M1690-02	1	Water	NA	NA	20.4	11.30	03/16/2021	17:12
7	M1690-02DUP	1	Water	NA	NA	20.5	11.31	03/16/2021	17:13
8	CCV2	1	Water	NA	NA	20.1	12.02	03/16/2021	17:15

WORKLIST(Hardcopy Internal Chain)

70 113538

WorkList Name : CORROSIONITY M1690 WorkList ID : 147189 Department : Wet-Chemistry Date : 03-16-2021 16:48:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/16/2021	Water	M1690-02 ~A	Corrosivity	Cool 4 deg C	DEWB01	H62	MC-WC	03/16/2021	9040C

Date/Time 03.16.21 16:50
 Received by: JD (wet Chem)
 Relinquished by: KS (wet Chem)

Date/Time 03.16.21 17:50
 Received by: KS (wet Chem)
 Relinquished by: JD (wet Chem)