

LB113548


INSTRUMENT ID: CV2

7471B

Linear

μ Abs.:

54691



A= 0.0000e+000

B= 1.8292e-004

C= -4.4257e-002

Rho= 0.9999516

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.037	-0.037	39	0.000	39					-
0.2	0.200	0.180	-0.020	1224	0.0 %	1224					10
2.5	2.500	2.540	0.040	14128	0.0 %	14128					2
5.0	5.000	5.054	0.054	27870	0.0 %	27870					1
7.5	7.500	7.504	0.004	41262	0.0 %	41262					0
10.0	10.000	9.960	-0.040	54691	0.0 %	54691					0

LB113548 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 17 Mar 2021 14:23:24					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	39	-	0.0000	7471B	PPB	17 Mar 2021 15:02:44	S	Std
0.2 - 1	502	1224	-	0.2000	7471B	PPB	17 Mar 2021 15:05:01	S	Std
2.5 - 1	525	14128	-	2.5000	7471B	PPB	17 Mar 2021 15:07:17	S	Std
5.0 - 1	55	27870	-	5.0000	7471B	PPB	17 Mar 2021 15:09:34	S	Std
7.5 - 1	57.5	41262	-	7.5000	7471B	PPB	17 Mar 2021 15:11:50	S	Std
10.0 - 1	510	54691	-	10.0000	7471B	PPB	17 Mar 2021 15:14:06	S	Std
ICV40 - 1	ICV40	20566	3.7178		-7471B	PPB	17 Mar 2021 15:16:57	U	SMPL
ICB40 - 1	ICB40	-32	-0.0501		-7471B	PPB	17 Mar 2021 15:19:12	U	SMPL
CCV35 - 1	CCV35	27900	5.0593		-7471B	PPB	17 Mar 2021 15:24:21	U	SMPL
CCB35 - 1	CCB35	-35	-0.0507		-7471B	PPB	17 Mar 2021 15:26:36	U	SMPL
CRA - 1	CRA	1262	0.1866		-7471B	PPB	17 Mar 2021 15:28:51	U	SMPL
HighStd - 1	HighStd	55038	10.0235		-7471B	PPB	17 Mar 2021 15:31:08	U	SMPL
ChkStd - 1	ChkStd	39380	7.1593		-7471B	PPB	17 Mar 2021 15:33:25	U	SMPL
PB135209BL - 1	PBS	-71	-0.0572		-7471B	PPB	17 Mar 2021 15:35:41	U	SMPL
PB135209BS - 1	LCSS	19755	3.5694		-7471B	PPB	17 Mar 2021 15:37:59	U	SMPL
M1553-13 - 1	347	6232	1.0957		-7471B	PPB	17 Mar 2021 15:40:15	U	SMPL
M1686-01 - 1	SW-VAN-WINKLE-STREET	1085	0.1542		-7471B	PPB	17 Mar 2021 15:42:31	U	SMPL
M1686-02 - 1	SW-SUMMER-STREET	3698	0.6322		-7471B	PPB	17 Mar 2021 15:44:48	U	SMPL
M1688-01 - 1	R-LA-4-1-WC	4632	0.8030		-7471B	PPB	17 Mar 2021 15:47:05	U	SMPL
M1689-01 - 1	R-LA-4-4-WC	12237	2.1942		-7471B	PPB	17 Mar 2021 15:49:20	U	SMPL
CCV36 - 1	CCV36	28300	5.1325		-7471B	PPB	17 Mar 2021 15:51:35	U	SMPL
CCB36 - 1	CCB36	-53	-0.0540		-7471B	PPB	17 Mar 2021 15:53:51	U	SMPL
M1691-01 - 1	FK-GF-02-068-A	6236	1.0965		-7471B	PPB	17 Mar 2021 15:56:05	U	SMPL
M1691-03 - 1	FK-GF-02-068-B	7292	1.2896		-7471B	PPB	17 Mar 2021 15:58:21	U	SMPL
M1692-01 - 1	FK-GF-02-069-A	6999	1.2360		-7471B	PPB	17 Mar 2021 16:00:36	U	SMPL
M1692-03 - 1	FK-GF-02-069-B	6936	1.2245		-7471B	PPB	17 Mar 2021 16:02:52	U	SMPL
M1693-01 - 1	TP-5A	5367	0.9375		-7471B	PPB	17 Mar 2021 16:05:07	U	SMPL
M1693-01DUP - 1	TP-5ADUP	5449	0.9525		-7471B	PPB	17 Mar 2021 16:10:05	U	SMPL
M1693-01MS - 1	TP-5AMS	27561	4.9973		-7471B	PPB	17 Mar 2021 16:12:21	U	SMPL
M1693-01MSD - 1	TP-5AMSD	26931	4.8821		-7471B	PPB	17 Mar 2021 16:14:38	U	SMPL
M1693-01LX5 - 1		1005	0.1396		-7471B	PPB	17 Mar 2021 16:16:55	U	SMPL
M1693-01A - 1		27001	4.8949		-7471B	PPB	17 Mar 2021 16:19:10	U	SMPL
CCV37 - 1	CCV37	28545	5.1773		-7471B	PPB	17 Mar 2021 16:21:25	U	SMPL
CCB37 - 1	CCB37	-36	-0.0508		-7471B	PPB	17 Mar 2021 16:23:41	U	SMPL