



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113559

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	03/17/2021	17:00
2	CAL2	1	Water	NA	NA	20.4	7.00	03/17/2021	17:01
3	CAL3	1	Water	NA	NA	20.4	10.03	03/17/2021	17:02
4	ICV	1	Water	NA	NA	20.4	6.98	03/17/2021	17:03
5	CCV1	1	Water	NA	NA	20.4	2.01	03/17/2021	17:05
6	M1712-01	1	Solid	20.03	20	20.2	9.43	03/17/2021	17:40
7	M1712-01DUP	1	Solid	20.04	20	20.3	9.44	03/17/2021	17:41
8	M1712-03	1	Solid	20.05	20	20.4	9.73	03/17/2021	17:42
9	M1712-05	1	Solid	20.01	20	20.2	9.46	03/17/2021	17:45
10	M1712-07	1	Solid	20.03	20	20.1	9.40	03/17/2021	17:46
11	M1712-09	1	Solid	20.04	20	20.3	10.12	03/17/2021	17:48
12	M1712-11	1	Solid	20.01	20	20.3	10.11	03/17/2021	18:00
13	M1712-13	1	Solid	20.03	20	20.4	9.30	03/17/2021	18:02
14	M1712-15	1	Solid	20.04	20	20.5	9.06	03/17/2021	18:05
15	M1712-17	1	Solid	20.04	20	20.6	9.80	03/17/2021	17:20
16	CCV2	1	Water	NA	NA	20.2	12.02	03/17/2021	17:21
17	M1712-19	1	Solid	20.02	20	20.7	9.29	03/17/2021	17:22
18	M1712-21	1	Solid	20.01	20	20.6	10.44	03/17/2021	17:25
19	M1712-23	1	Solid	20.03	20	20.5	9.40	03/17/2021	17:30
20	CCV3	1	Water	NA	NA	20.3	2.01	03/17/2021	17:31

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M1712 WorkList ID : 147238 Department : Wet-Chemistry Date : 03-17-2021 15:42:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/17/2021	Solid	M1712-01	pH	Cool 4 deg C	PSEG03	I13	EXTERNAL-PAD-1	03/17/2021	9045D
03/17/2021	Solid	M1712-03	pH	Cool 4 deg C	PSEG03	I13	EXTERNAL-PAD-2	03/17/2021	9045D
03/17/2021	Solid	M1712-05	pH	Cool 4 deg C	PSEG03	I13	RED-BRICK-1	03/17/2021	9045D
03/17/2021	Solid	M1712-07	pH	Cool 4 deg C	PSEG03	I13	RED-BRICK-2	03/17/2021	9045D
03/17/2021	Solid	M1712-09	pH	Cool 4 deg C	PSEG03	I13	NEW-BLOCK	03/17/2021	9045D
03/17/2021	Solid	M1712-11	pH	Cool 4 deg C	PSEG03	I13	FLOOR-PAT	03/17/2021	9045D
03/17/2021	Solid	M1712-13	pH	Cool 4 deg C	PSEG03	I13	YELLOW-WALL-1	03/17/2021	9045D
03/17/2021	Solid	M1712-15	pH	Cool 4 deg C	PSEG03	I13	YELLOW-WALL-2	03/17/2021	9045D
03/17/2021	Solid	M1712-17	pH	Cool 4 deg C	PSEG03	I13	EXTERNAL-PAD-COMP	03/17/2021	9045D
03/17/2021	Solid	M1712-19	pH	Cool 4 deg C	PSEG03	I13	RED-BRICK-COMP	03/17/2021	9045D
03/17/2021	Solid	M1712-21	pH	Cool 4 deg C	PSEG03	I13	NEW-BLOCK-FLOOR-PAT	03/17/2021	9045D
03/17/2021	Solid	M1712-23	pH	Cool 4 deg C	PSEG03	I13	YELLOW-WALL-COMP	03/17/2021	9045D

Date/Time 03-17-21 16:00:00

Received by: JS (Wetchem)

Relinquished by: JS (Wetchem)

Date/Time 03-17-21 18:00:00

Received by: JS (Wetchem)

Relinquished by: JS (Wetchem)