

LB113561


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41112



A= 0.0000e+000

B= 2.4173e-004

C= -2.2100e-002

Rho= 0.9997547

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/b1)
0.0	0.000	-0.014	-0.014	33	0.000	33					—
0.2	0.200	0.176	-0.024	821	0.0 %	821					-12
2.5	2.500	2.565	0.065	10702	0.0 %	10702					3
5.0	5.000	4.914	-0.086	20418	0.0 %	20418					-2
7.5	7.500	7.643	0.143	31711	0.0 %	31711					2
10.0	10.000	9.916	-0.084	41112	0.0 %	41112					-1

LB113561 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Mar 2021 11:16:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	33	-	0.0000	7470A	PPB	18 Mar 2021 11:34:32	S	Std
0.2 - 1	50.2	821	-	0.2000	7470A	PPB	18 Mar 2021 11:36:49	S	Std
2.5 - 1	52.5	10702	-	2.5000	7470A	PPB	18 Mar 2021 11:39:06	S	Std
5.0 - 1	55	20418	-	5.0000	7470A	PPB	18 Mar 2021 11:41:23	S	Std
7.5 - 1	59.5	31711	-	7.5000	7470A	PPB	18 Mar 2021 11:43:39	S	Std
10.0 - 1	510	41112	-	10.0000	7470A	PPB	18 Mar 2021 11:45:54	S	Std
ICV41 - 1	ICV41	15420	3.7054	-	7470A	PPB	18 Mar 2021 11:48:45	U	SMPL
ICB41 - 1	ICB41	97	0.0013	-	7470A	PPB	18 Mar 2021 11:53:29	U	SMPL
CCV38 - 1	CCV38	20402	4.9097	-	7470A	PPB	18 Mar 2021 11:55:43	U	SMPL
CCB38 - 1	CCB38	79	-0.0030	-	7470A	PPB	18 Mar 2021 11:57:58	U	SMPL
CRA - 1	CRA	815	0.1749	-	7470A	PPB	18 Mar 2021 12:00:13	U	SMPL
HighStd - 1	HighStd	41554	10.0228	-	7470A	PPB	18 Mar 2021 12:02:27	U	SMPL
ChkStd - 1	ChkStd	30000	7.2298	-	7470A	PPB	18 Mar 2021 12:04:42	U	SMPL
PB135210BL - 1	PBW	92	0.0001	-	7470A	PPB	18 Mar 2021 12:06:58	U	SMPL
PB135210BS - 1	LCSW	15710	3.7755	-	7470A	PPB	18 Mar 2021 12:09:14	U	SMPL
M1687-29 - 1	AUD-1031-1032	135	0.0105	-	7470A	PPB	18 Mar 2021 12:11:30	U	SMPL
M1687-29DUP - 1	AUD-1031-1032DUP	169	0.0188	-	7470A	PPB	18 Mar 2021 12:13:46	U	SMPL
M1687-29MS - 1	AUD-1031-1032MS	15726	3.7793	-	7470A	PPB	18 Mar 2021 12:16:05	U	SMPL
M1687-29MSD - 1	AUD-1031-1032MSD	15932	3.8291	-	7470A	PPB	18 Mar 2021 12:18:23	U	SMPL
M1687-33 - 1	AUD-1044-1046	206	0.0277	-	7470A	PPB	18 Mar 2021 12:20:41	U	SMPL
CCV39 - 1	CCV39	20535	4.9418	-	7470A	PPB	18 Mar 2021 12:22:55	U	SMPL
CCB39 - 1	CCB39	117	0.0062	-	7470A	PPB	18 Mar 2021 12:27:55	U	SMPL
M1687-38 - 1	AUD-1049-1051	304	0.0514	-	7470A	PPB	18 Mar 2021 12:30:10	U	SMPL
M1687-43 - 1	AUD-1055	362	0.0654	-	7470A	PPB	18 Mar 2021 12:32:26	U	SMPL
M1690-02 - 1	MC-WC	344	0.0611	-	7470A	PPB	18 Mar 2021 12:34:41	U	SMPL
PB135211BL - 1	PBW	314	0.0538	-	7470A	PPB	18 Mar 2021 12:36:57	U	SMPL
PB135211BS - 1	LCSW	15493	3.7230	-	7470A	PPB	18 Mar 2021 12:39:12	U	SMPL
M1687-05 - 1	AUD-1022	191	0.0241	-	7470A	PPB	18 Mar 2021 12:41:28	U	SMPL
M1687-08 - 1	AUD-1033-1042	3151	0.7396	-	7470A	PPB	18 Mar 2021 12:43:44	U	SMPL
M1687-08DUP - 1	AUD-1033-1042DUP	2796	0.6538	-	7470A	PPB	18 Mar 2021 12:46:01	U	SMPL
M1687-08MS - 1	AUD-1033-1042MS	19766	4.7559	-	7470A	PPB	18 Mar 2021 12:48:18	U	SMPL
M1687-08MSD - 1	AUD-1033-1042MSD	18565	4.4656	-	7470A	PPB	18 Mar 2021 12:50:34	U	SMPL
CCV40 - 1	CCV40	20394	4.9077	-	7470A	PPB	18 Mar 2021 12:52:49	U	SMPL
CCB40 - 1	CCB40	-6	-0.0236	-	7470A	PPB	18 Mar 2021 12:55:04	U	SMPL
M1687-20 - 1	AUD-1047	3763	0.8875	-	7470A	PPB	18 Mar 2021 12:57:20	U	SMPL
M1687-23 - 1	AUD-1048	349	0.0623	-	7470A	PPB	18 Mar 2021 12:59:35	U	SMPL
M1687-26 - 1	AUD-1056	343	0.0608	-	7470A	PPB	18 Mar 2021 13:01:51	U	SMPL
M1690-01 - 1	MC-WC	1257	0.2818	-	7470A	PPB	18 Mar 2021 13:04:06	U	SMPL
PB135210TB - 1	PB135188TB	207	0.0279	-	7470A	PPB	18 Mar 2021 13:06:22	U	SMPL
M1687-29LX5 - 1		118	0.0064	-	7470A	PPB	18 Mar 2021 13:08:38	U	SMPL
M1687-29A - 1		15787	3.7941	-	7470A	PPB	18 Mar 2021 13:10:54	U	SMPL
M1687-08LX5 - 1		584	0.1191	-	7470A	PPB	18 Mar 2021 13:13:11	U	SMPL
M1687-08A - 1		19022	4.5761	-	7470A	PPB	18 Mar 2021 13:15:28	U	SMPL
CCV41 - 1	CCV41	20579	4.9525	-	7470A	PPB	18 Mar 2021 13:17:43	U	SMPL
CCB41 - 1	CCB41	44	-0.0115	-	7470A	PPB	18 Mar 2021 13:19:59	U	SMPL