


LB113568

7471B

INSTRUMENT ID: CV1

 μ Abs.:

55975



Linear

A= 0.0000e+000

B= 1.7647e-004

C= -1.4184e-002

Rho= 0.9996675

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Std
0.0	0.000	0.002	0.002	91	0.000	91					-
0.2	0.200	0.193	-0.007	1176	0.0 %	1176					-4
2.5	2.500	2.423	-0.077	13813	0.0 %	13813					-3
5.0	5.000	5.053	0.053	28711	0.0 %	28711					1
7.5	7.500	7.665	0.165	43513	0.0 %	43513					2
10.0	10.000	9.864	-0.136	55975	0.0 %	55975					-1

LB113568 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Mar 2021 14:48:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	91	-	0.0000	7471B	PPB	18 Mar 2021 14:51:29	S	Std
0.2 - 1	50.2	1176	-	0.2000	7471B	PPB	18 Mar 2021 14:53:46	S	Std
2.5 - 1	52.5	13813	-	2.5000	7471B	PPB	18 Mar 2021 14:56:03	S	Std
5.0 - 1	55	28711	-	5.0000	7471B	PPB	18 Mar 2021 14:58:20	S	Std
7.5 - 1	57.5	43513	-	7.5000	7471B	PPB	18 Mar 2021 15:00:36	S	Std
10.0 - 1	510	55975	-	10.0000	7471B	PPB	18 Mar 2021 15:02:52	S	Std
ICV42 - 1	ICV42	21920	3.8541	-	7471B	PPB	18 Mar 2021 15:05:36	U	SMPL
ICB42 - 1	ICB42	32	-0.0085	-	7471B	PPB	18 Mar 2021 15:07:50	U	SMPL
CCV42 - 1	CCV42	28545	5.0233	-	7471B	PPB	18 Mar 2021 15:10:06	U	SMPL
CCB42 - 1	CCB42	18	-0.0110	-	7471B	PPB	18 Mar 2021 15:12:21	U	SMPL
CRA - 1	CRA	1172	0.1926	-	7471B	PPB	18 Mar 2021 15:14:36	U	SMPL
HighStd - 1	HighStd	55254	9.7367	-	7471B	PPB	18 Mar 2021 15:16:51	U	SMPL
ChkStd - 1	ChkStd	40096	7.0617	-	7471B	PPB	18 Mar 2021 15:19:06	U	SMPL
PB135238BL - 1	PBS	-30	-0.0195	-	7471B	PPB	18 Mar 2021 15:21:23	U	SMPL
PB135238BS - 1	LCSS	21372	3.7574	-	7471B	PPB	18 Mar 2021 15:23:39	U	SMPL
M1678-03 - 1	D2470-D2800(SOIL)-B	10627	1.8612	-	7471B	PPB	18 Mar 2021 15:25:55	U	SMPL
M1682-02 - 1	SP-MANHOLE	12819	2.2480	-	7471B	PPB	18 Mar 2021 15:28:11	U	SMPL
M1706-02 - 1	COMP-P-1:P-2:P-3:P-4	45	-0.0062	-	7471B	PPB	18 Mar 2021 15:30:28	U	SMPL
M1707-02 - 1	ELEVATOR-510-EL-167-157	1806	0.3045	-	7471B	PPB	18 Mar 2021 15:32:45	U	SMPL
M1707-05 - 1	ELEVATOR-511-EL-167-162	654	0.1012	-	7471B	PPB	18 Mar 2021 15:35:00	U	SMPL
CCV43 - 1	CCV43	28838	5.0750	-	7471B	PPB	18 Mar 2021 15:37:15	U	SMPL
CCB43 - 1	CCB43	30	-0.0089	-	7471B	PPB	18 Mar 2021 15:39:30	U	SMPL
M1707-08 - 1	ELEVATOR-511-EL-162-157	146	0.0116	-	7471B	PPB	18 Mar 2021 15:41:45	U	SMPL
M1709-02 - 1	PSHQ-SOIL	7650	1.3358	-	7471B	PPB	18 Mar 2021 15:44:00	U	SMPL
M1713-01 - 1	TP-6A	14087	2.4718	-	7471B	PPB	18 Mar 2021 15:46:16	U	SMPL
M1713-01DUP - 1	TP-6ADUP	14138	2.4808	-	7471B	PPB	18 Mar 2021 15:48:31	U	SMPL
M1713-01MS - 1	TP-6AMS	37041	6.5226	-	7471B	PPB	18 Mar 2021 15:53:38	U	SMPL
M1713-01MSD - 1	TP-6AMSD	34664	6.1031	-	7471B	PPB	18 Mar 2021 15:55:54	U	SMPL
M1713-03 - 1	TP-7A	14980	2.6294	-	7471B	PPB	18 Mar 2021 15:58:10	U	SMPL
M1713-01LX5 - 1		2773	0.4752	-	7471B	PPB	18 Mar 2021 16:00:28	U	SMPL
M1713-01A - 1		39228	6.9086	-	7471B	PPB	18 Mar 2021 16:02:45	U	SMPL
CCV44 - 1	CCV44	29014	5.1060	-	7471B	PPB	18 Mar 2021 16:05:00	U	SMPL
CCB44 - 1	CCB44	32	-0.0085	-	7471B	PPB	18 Mar 2021 16:07:16	U	SMPL