



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113572

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	03/18/2021	12:25
2	M1690-02		1	10	100.00	>212.0	>212.0	03/18/2021	12:50
3	M1690-02DUP		1	10	100.00	>212.0	>212.0	03/18/2021	13:15
4	M1702-08		1	11	100.00	>212.0	>212.0	03/18/2021	13:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB113572

WorkList Name : FP31721-1 WorkList ID : 147218 Department : Wet-Chemistry Date : 03-17-2021 11:46:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/19/2021	Water	M1690-02	Flash Point	Cool 4 deg C	DEWB01	H62	MC-WC	03/16/2021	1010A
03/24/2021	Water	M1702-08	Flash Point	Cool 4 deg C	PSEG03	I11	33021	03/17/2021	1010A

Date/Time 03/18/21 12:00
 Received by: KS (Wetchem)
 Relinquished by: SP (Wetchem)

Date/Time 03/18/21 13:45
 Received by: SP (Wetchem)
 Relinquished by: KS (Wetchem)