



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113575

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	3.99	03/18/2021	15:10
2	CAL2	1	Water	NA	NA	20.2	7.00	03/18/2021	15:11
3	CAL3	1	Water	NA	NA	20.2	10.03	03/18/2021	15:12
4	ICV	1	Water	NA	NA	20.2	6.99	03/18/2021	15:15
5	CCV1	1	Water	NA	NA	20.2	2.01	03/18/2021	15:16
6	M1702-08	1	Water	NA	NA	20.1	6.34	03/18/2021	15:17
7	M1702-08DUP	1	Water	NA	NA	20.2	6.35	03/18/2021	15:18
8	M1717-01	1	Water	NA	NA	20.4	5.09	03/18/2021	15:20
9	M1724-04	1	Water	NA	NA	20.7	6.72	03/18/2021	15:21
10	CCV2	1	Water	NA	NA	20.2	12.02	03/18/2021	15:25

LB 113575

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m1724 WorkList ID : 147259 Department : Wet-Chemistry Date : 03-18-2021 14:59:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/17/2021	Water	M1702-08	pH	Cool 4 deg C	PSEG03	I11	33021	03/17/2021	9040C
03/15/2021	Water	M1717-01	pH	Cool 4 deg C	STER14	J61	JS-SYSTEM	03/15/2021	9040C
03/18/2021	Water	M1724-04	pH	Cool 4 deg C	PSEG03	I21	P1679	03/18/2021	9040C

Date/Time 03/18/21 15:00
Received by: JP (wet-chem)
Relinquished by: JP (wet-chem)

Date/Time 03-18-21 16:10
Received by: JP (wet-chem)
Relinquished by: JP (wet-chem)