



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113588

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	3.99	03/19/2021	10:10
2	CAL2	1	Water	NA	NA	20.4	7.00	03/19/2021	10:11
3	CAL3	1	Water	NA	NA	20.4	10.03	03/19/2021	10:12
4	ICV	1	Water	NA	NA	20.4	6.99	03/19/2021	10:15
5	CCV1	1	Water	NA	NA	20.4	2.01	03/19/2021	10:16
6	M1725-01	1	Solid	20.03	20	21.1	6.21	03/19/2021	10:19
7	M1725-01DUP	1	Solid	20.04	20	21.2	6.22	03/19/2021	10:20
8	CCV2	1	Water	NA	NA	20.4	12.02	03/19/2021	10:21

WORKLIST(Hardcopy Internal Chain)

VB 113588

WorkList Name : ph m1725

WorkList ID : 147292

Department : Wet-Chemistry

Date : 03-19-2021 10:00:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/18/2021	Solid	M1725-01	pH	Cool 4 deg C	PSEG03	122	72-11948-124018	03/18/2021	9045D

Date/Time 03-19-21 10:05
 Received by: JN Wet Chem
 Relinquished by: JN (Sh)

Date/Time 03-19-21 11:01
 Received by: JR (S MS)
 Relinquished by: JR Wet Chem