



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB113593

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	03/19/2021	12:35
2	CAL2	1	Water	NA	NA	20.3	7.00	03/19/2021	12:36
3	CAL3	1	Water	NA	NA	20.3	10.03	03/19/2021	12:37
4	ICV	1	Water	NA	NA	20.3	6.98	03/19/2021	12:40
5	CCV1	1	Water	NA	NA	20.3	2.01	03/19/2021	12:41
6	M1742-01	1	Water	NA	NA	20.7	6.73	03/19/2021	12:42
7	M1742-01DUP	1	Water	NA	NA	20.6	6.74	03/19/2021	12:43
8	CCV2	1	Water	NA	NA	20.4	12.02	03/19/2021	12:45

WORKLIST(Hardcopy Internal Chain)

18 113543

WorkList Name : corrosivity 11741 water WorkList ID : 147297 Department : Wet-Chemistry Date : 03-19-2021 12:25:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/18/2021	Water	M1742-01	Corrosivity	Cool 4 deg C	TETR06	J11	CAL-IDWTANK1-20210318	03/18/2021	9040C

Date/Time 03-19-21 12:30
 Received by: JO Clem Chem
 Relinquished by: AM Cassel Chem

Date/Time 03-19-21 13:33
 Received by: AM (wet-chem)
 Relinquished by: JO (wet) Chem