



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113596

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	03/19/2021	09:25
2	M1687-49		1	8	78.00	172.4	172.1	03/19/2021	09:50
3	M1687-49DUP		1	8	80.00	176.0	175.7	03/19/2021	10:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 113596

WorkList Name : FP31721

WorkList ID : 147217

Department : Wet-Chemistry

Date : 03-17-2021 11:45:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/29/2021	Solid	M1687-49 B	Flash Point	Cool 4 deg C	PSEG03	H61	31144	03/15/2021	1010A

Date/Time 03/19/21 09:00
 Received by: KS (Wetchem)
 Relinquished by: JP (Wet Chem)

Date/Time 03/19/21 10:25
 Received by: KS (Wetchem)
 Relinquished by: JP (Wetchem)