

LA113598


7471B

INSTRUMENT ID: CY1

Linear

μ Abs.:

56230



A= 0.0000e+000

B= 1.7833e-004

C= 1.0955e-002

Rho= 0.9999542

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.028	0.028	97	0.000	97					
0.2	0.200	0.214	0.014	1136	0.0 %	1136					
2.5	2.500	2.494	-0.006	13923	0.0 %	13923					
5.0	5.000	4.929	-0.071	27581	0.0 %	27581					
7.5	7.500	7.497	-0.003	41978	0.0 %	41978					
10.0	10.000	10.038	0.038	56230	0.0 %	56230					

0/00
1
7
0
1
0
0
0

LB113598 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Mar 2021 14:29:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	97	-	0.0000	7471B	PPB	19 Mar 2021 14:33:28	S	Std
0.2 - 1	502	1136	-	0.2000	7471B	PPB	19 Mar 2021 14:35:45	S	Std
2.5 - 1	52.5	13923	-	2.5000	7471B	PPB	19 Mar 2021 14:38:03	S	Std
5.0 - 1	55	27581	-	5.0000	7471B	PPB	19 Mar 2021 14:40:19	S	Std
7.5 - 1	57.5	41978	-	7.5000	7471B	PPB	19 Mar 2021 14:42:35	S	Std
10.0 - 1	58	56230	-	10.0000	7471B	PPB	19 Mar 2021 14:44:51	S	Std
ICV44 - 1	ICV44	22106	3.9530	-	7471B	PPB	19 Mar 2021 14:47:34	U	SMPL
ICB44 - 1	ICB44	12	0.0131	-	7471B	PPB	19 Mar 2021 14:49:48	U	SMPL
CCV49 - 1	CCV49	27849	4.9772	-	7471B	PPB	19 Mar 2021 14:52:03	U	SMPL
CCB49 - 1	CCB49	4	0.0117	-	7471B	PPB	19 Mar 2021 14:54:17	U	SMPL
CRA - 1	CRA	1125	0.2116	-	7471B	PPB	19 Mar 2021 14:56:32	U	SMPL
HighStd - 1	HighStd	56605	10.1051	-	7471B	PPB	19 Mar 2021 14:58:47	U	SMPL
ChkStd - 1	ChkStd	39737	7.0971	-	7471B	PPB	19 Mar 2021 15:01:02	U	SMPL
PB135253BL - 1	PBS	-3	0.0104	-	7471B	PPB	19 Mar 2021 15:03:19	U	SMPL
PB135253BS - 1	LCSS	20414	3.6513	-	7471B	PPB	19 Mar 2021 15:05:37	U	SMPL
M1718-01 - 1	BACKFILL-ASHCROFT-03	408	0.0837	-	7471B	PPB	19 Mar 2021 15:07:53	U	SMPL
M1718-01DUP - 1	BACKFILL-ASHCROFT-03DUP	226	0.0513	-	7471B	PPB	19 Mar 2021 15:10:09	U	SMPL
M1718-01MS - 1	BACKFILL-ASHCROFT-03MS	20987	3.7535	-	7471B	PPB	19 Mar 2021 15:14:57	U	SMPL
M1718-01MSD - 1	BACKFILL-ASHCROFT-03MSD	21518	3.8482	-	7471B	PPB	19 Mar 2021 15:17:13	U	SMPL
M1718-02 - 1	BACKFILL-ASHCROFT-04	48	0.0195	-	7471B	PPB	19 Mar 2021 15:19:28	U	SMPL
CCV50 - 1	CCV50	27918	4.9895	-	7471B	PPB	19 Mar 2021 15:21:43	U	SMPL
CCB50 - 1	CCB50	25	0.0154	-	7471B	PPB	19 Mar 2021 15:23:59	U	SMPL
M1718-03 - 1	BACKFILL-ASHCROFT-05	132	0.0345	-	7471B	PPB	19 Mar 2021 15:26:14	U	SMPL
M1721-01 - 1	TR-05-031821	6747	1.2141	-	7471B	PPB	19 Mar 2021 15:28:30	U	SMPL
M1721-03 - 1	TR-08-031821	901	0.1716	-	7471B	PPB	19 Mar 2021 15:30:45	U	SMPL
M1722-01 - 1	HD-01-031821	1741	0.3214	-	7471B	PPB	19 Mar 2021 15:33:00	U	SMPL
M1723-01 - 1	24055	2323	0.4252	-	7471B	PPB	19 Mar 2021 15:35:16	U	SMPL
M1723-03 - 1	RO124027	1617	0.2993	-	7471B	PPB	19 Mar 2021 15:37:32	U	SMPL
M1724-01 - 1	26207	126	0.0334	-	7471B	PPB	19 Mar 2021 15:39:48	U	SMPL
M1718-01LX5 - 1		83	0.0258	-	7471B	PPB	19 Mar 2021 15:42:05	U	SMPL
M1718-01A - 1		23185	4.1455	-	7471B	PPB	19 Mar 2021 15:44:22	U	SMPL
CCV51 - 1	CCV51	27896	4.9856	-	7471B	PPB	19 Mar 2021 15:46:37	U	SMPL
CCB51 - 1	CCB51	8	0.0124	-	7471B	PPB	19 Mar 2021 15:48:52	U	SMPL