



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113612

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	3.99	03/22/2021	08:00
2	CAL2	1	Water	NA	NA	20.2	7.00	03/22/2021	08:01
3	CAL3	1	Water	NA	NA	20.2	10.02	03/22/2021	08:02
4	ICV	1	Water	NA	NA	20.2	6.99	03/22/2021	08:05
5	CCV1	1	Water	NA	NA	20.2	2.01	03/22/2021	08:06
6	M1750-02	1	Water	NA	NA	13.4	7.51	03/22/2021	08:08
7	M1752-02	1	Water	NA	NA	20.5	6.58	03/22/2021	08:09
8	M1752-03	1	Water	NA	NA	20.1	4.89	03/22/2021	08:10
9	M1752-04	1	Water	NA	NA	20.3	6.72	03/22/2021	08:11
10	M1752-05	1	Water	NA	NA	20.4	5.15	03/22/2021	08:12
11	M1752-06	1	Water	NA	NA	20.3	5.89	03/22/2021	08:15
12	M1752-08	1	Water	NA	NA	20.1	5.99	03/22/2021	08:16
13	M1752-09	1	Water	NA	NA	20.3	6.67	03/22/2021	08:17
14	M1752-10	1	Water	NA	NA	20.1	6.70	03/22/2021	08:18
15	M1752-11	1	Water	NA	NA	20.1	5.46	03/22/2021	08:20
16	CCV2	1	Water	NA	NA	20.3	12.02	03/22/2021	08:22
17	M1752-12	1	Water	NA	NA	20.3	6.72	03/22/2021	08:23
18	M1752-13	1	Water	NA	NA	20.3	7.17	03/22/2021	08:25
19	M1753-01	1	Water	NA	NA	20.1	6.69	03/22/2021	08:26
20	M1753-03	1	Water	NA	NA	20.3	7.64	03/22/2021	08:28

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	M1754-01	1	Water	NA	NA	20.1	7.20	03/22/2021	08:30
22	M1754-01DUP	1	Water	NA	NA	20.2	7.21	03/22/2021	08:31
23	CCV3	1	Water	NA	NA	20.3	2.01	03/22/2021	08:32

WORKLIST(Hardcopy Internal Chain)

18113612

WorkList Name : ph m1750

WorkList ID : 147319

Department : Wet-Chemistry

Date : 03-22-2021 07:46:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/19/2021	Water	M1750-02	pH	Cool 4 deg C	PSEG03	I22	TRE-1015	03/19/2021	9040C
03/19/2021	Water	M1752-02	pH	Cool 4 deg C	PSEG03	I23	41711	03/19/2021	9040C
03/19/2021	Water	M1752-03	pH	Cool 4 deg C	PSEG03	I23	41714	03/19/2021	9040C
03/19/2021	Water	M1752-04	pH	Cool 4 deg C	PSEG03	I23	41715	03/19/2021	9040C
03/19/2021	Water	M1752-05	pH	Cool 4 deg C	PSEG03	I23	41716	03/19/2021	9040C
03/19/2021	Water	M1752-06	pH	Cool 4 deg C	PSEG03	I23	41717	03/19/2021	9040C
03/19/2021	Water	M1752-08	pH	Cool 4 deg C	PSEG03	I23	41720	03/19/2021	9040C
03/19/2021	Water	M1752-09	pH	Cool 4 deg C	PSEG03	I23	41721	03/19/2021	9040C
03/19/2021	Water	M1752-10	pH	Cool 4 deg C	PSEG03	I23	41722	03/19/2021	9040C
03/19/2021	Water	M1752-11	pH	Cool 4 deg C	PSEG03	I23	41723	03/19/2021	9040C
03/19/2021	Water	M1752-12	pH	Cool 4 deg C	PSEG03	I23	41718	03/19/2021	9040C
03/19/2021	Water	M1752-13	pH	Cool 4 deg C	PSEG03	I23	30816	03/19/2021	9040C
03/19/2021	Water	M1753-01	pH	Cool 4 deg C	PSEG03	I23	33027	03/19/2021	9040C
03/19/2021	Water	M1753-03	pH	Cool 4 deg C	PSEG03	I23	32539	03/19/2021	9040C
03/19/2021	Water	M1754-01	pH	Cool 4 deg C	PSEG03	I23	N48987	03/19/2021	9040C

Date/Time 03/22/21 7:50
 Received by: JD (wet chem)
 Relinquished by: SP (wet chem)

Date/Time 03/22/21 9:10
 Received by: SP (wet chem)
 Relinquished by: JD (wet chem)