



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113613

Slope : 97.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	3.99	03/22/2021	08:00
2	CAL2	1	Water	NA	NA	20.2	7.00	03/22/2021	08:01
3	CAL3	1	Water	NA	NA	20.2	10.02	03/22/2021	08:02
4	ICV	1	Water	NA	NA	20.2	6.99	03/22/2021	08:05
5	CCV1	1	Water	NA	NA	20.2	2.01	03/22/2021	08:06
6	M1751-01	1	Solid	20.03	20	20.3	7.05	03/22/2021	08:35
7	M1751-01DUP	1	Solid	20.04	20	20.4	7.06	03/22/2021	08:36
8	CCV2	1	Water	NA	NA	20.3	12.02	03/22/2021	08:37

WORKLIST(Hardcopy Internal Chain)

283 113613

WorkList Name : ph m1751 s

WorkList ID : 147320

Department : Wet-Chemistry

Date : 03-22-2021 07:47:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/19/2021	Solid	M1751-01	pH	Cool 4 deg C	PSEG05	I23	BU-03-031921	03/19/2021	9045D

Date/Time 03/22/21 7:50

Received by: JO West-Horn

Relinquished by: JO West-Horn

Date/Time 03/22/21 9:10

Received by: JO West-Horn

Relinquished by: JO West-Horn