


LB 113617

7470A

INSTRUMENT ID: CY2



Linear

A= 0.0000e+000

B= 2.4441e-004

C= -1.4213e-002

Rho= 0.9998084

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.004	0.004	75	0.000	75					-8
0.2	0.200	0.184	-0.016	810	0.0 %	810					-3
2.5	2.500	2.422	-0.078	9968	0.0 %	9968					2
5.0	5.000	5.116	0.116	20988	0.0 %	20988					1
7.5	7.500	7.562	0.062	30999	0.0 %	30999					-1
10.0	10.000	9.912	-0.088	40613	0.0 %	40613					

LB113617 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 22 Mar 2021 11:58:30					
Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	75	-	0.00007470A	PPB		22 Mar 2021 12:08:17	S	Std
0.2 - 1	50.2	810	-	0.20007470A	PPB		22 Mar 2021 12:10:34	S	Std
2.5 - 1	50.5	9968	-	2.50007470A	PPB		22 Mar 2021 12:12:51	S	Std
5.0 - 1	55	20988	-	5.00007470A	PPB		22 Mar 2021 12:15:08	S	Std
7.5 - 1	57.5	30999	-	7.50007470A	PPB		22 Mar 2021 12:19:57	S	Std
10.0 - 1	51V	40613	-	10.00007470A	PPB		22 Mar 2021 12:22:13	S	Std
ICV45 - 1	ICV45	15696	3.8221	-7470A	PPB		22 Mar 2021 12:24:55	U	SMPL
ICB45 - 1	ICB45	2	-0.0137	-7470A	PPB		22 Mar 2021 12:27:09	U	SMPL
CCV52 - 1	CCV52	20309	4.9496	-7470A	PPB		22 Mar 2021 12:29:24	U	SMPL
CCB52 - 1	CCB52	-8	-0.0162	-7470A	PPB		22 Mar 2021 12:31:39	U	SMPL
CRA - 1	CRA	799	0.1811	-7470A	PPB		22 Mar 2021 12:33:54	U	SMPL
HighStd - 1	HighStd	40918	9.9867	-7470A	PPB		22 Mar 2021 12:36:08	U	SMPL
ChkStd - 1	ChkStd	27942	6.8152	-7470A	PPB		22 Mar 2021 12:38:24	U	SMPL
PB135261BL - 1	PBW	-38	-0.0235	-7470A	PPB		22 Mar 2021 12:40:40	U	SMPL
PB135261BS - 1	LCSW	14862	3.6183	-7470A	PPB		22 Mar 2021 12:42:56	U	SMPL
M1718-04 - 1	BACKFILL-ASHCROFT-03	131	0.0178	-7470A	PPB		22 Mar 2021 12:50:01	U	SMPL
M1718-04DUP - 1	BACKFILL-ASHCROFT-03DUP	120	0.0151	-7470A	PPB		22 Mar 2021 12:52:17	U	SMPL
M1718-04MS - 1	BACKFILL-ASHCROFT-03MS	14425	3.5114	-7470A	PPB		22 Mar 2021 12:54:33	U	SMPL
M1718-04MSD - 1	BACKFILL-ASHCROFT-03MSD	15255	3.7143	-7470A	PPB		22 Mar 2021 12:56:50	U	SMPL
M1718-05 - 1	BACKFILL-ASHCROFT-04	61	0.0007	-7470A	PPB		22 Mar 2021 12:59:05	U	SMPL
CCV53 - 1	CCV53	20875	5.0879	-7470A	PPB		22 Mar 2021 13:01:20	U	SMPL
CCB53 - 1	CCB53	6	-0.0127	-7470A	PPB		22 Mar 2021 13:03:35	U	SMPL
M1718-06 - 1	BACKFILL-ASHCROFT-05	137	0.0193	-7470A	PPB		22 Mar 2021 13:05:50	U	SMPL
PB135262BL - 1	PBW	37	-0.0052	-7470A	PPB		22 Mar 2021 13:08:05	U	SMPL
PB135262BS - 1	LCSW	16037	3.9054	-7470A	PPB		22 Mar 2021 13:10:21	U	SMPL
M1724-04 - 1	P1679	831	0.1889	-7470A	PPB		22 Mar 2021 13:12:36	U	SMPL
M1724-04DUP - 1	P1679DUP	855	0.1948	-7470A	PPB		22 Mar 2021 13:14:52	U	SMPL
M1724-04MS - 1	P1679MS	17118	4.1697	-7470A	PPB		22 Mar 2021 13:17:09	U	SMPL
M1724-04MSD - 1	P1679MSD	16491	4.0164	-7470A	PPB		22 Mar 2021 13:19:25	U	SMPL
M1742-01 - 1	CAL-IDWTANK1-20210318	5	-0.0130	-7470A	PPB		22 Mar 2021 13:21:42	U	SMPL
PB135261TB - 1	PB135233TB	38	-0.0049	-7470A	PPB		22 Mar 2021 13:23:58	U	SMPL
CCV54 - 1	CCV54	20821	5.0747	-7470A	PPB		22 Mar 2021 13:26:14	U	SMPL
CCB54 - 1	CCB54	6	-0.0127	-7470A	PPB		22 Mar 2021 13:28:28	U	SMPL
M1718-04LX5 - 1		109	0.0124	-7470A	PPB		22 Mar 2021 13:30:44	U	SMPL
M1718-04A - 1		15695	3.8219	-7470A	PPB		22 Mar 2021 13:32:59	U	SMPL
M1724-04LX5 - 1		180	0.0298	-7470A	PPB		22 Mar 2021 13:35:15	U	SMPL
M1724-04A - 1		16591	4.0408	-7470A	PPB		22 Mar 2021 13:37:30	U	SMPL
CCV55 - 1	CCV55	20623	5.0263	-7470A	PPB		22 Mar 2021 13:39:46	U	SMPL
CCB55 - 1	CCB55	16	-0.0103	-7470A	PPB		22 Mar 2021 13:42:01	U	SMPL