



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113627

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/23/2021	08:10
2	CAL2	1	Water	NA	NA	20.3	7.00	03/23/2021	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	03/23/2021	08:12
4	ICV	1	Water	NA	NA	20.3	6.98	03/23/2021	08:15
5	CCV1	1	Water	NA	NA	20.3	2.01	03/23/2021	08:16
6	M1764-01	1	Solid	20.02	20	21.4	6.75	03/23/2021	08:30
7	M1764-01DUP	1	Solid	20.01	20	21.5	6.76	03/23/2021	08:31
8	CCV2	1	Water	NA	NA	20.3	12.02	03/23/2021	08:32

WORKLIST(Hardcopy Internal Chain)

13624

WorkList Name : PH M1764 WorkList ID : 147366 Department : Wet-Chemistry Date : 03-23-2021 08:02:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/22/2021	Solid	M1764-01	pH	Cool 4 deg C	PSEG03	J21	BUR-1009	03/22/2021	9045D

Date/Time 03/23/21 8:05
 Received by: JP (Webster)
 Relinquished by: KS (Webster)

Date/Time 03/23/21 10:00
 Received by: KS (Webster)
 Relinquished by: JP (Webster)