



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113628

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/23/2021	08:10
2	CAL2	1	Water	NA	NA	20.3	7.00	03/23/2021	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	03/23/2021	08:12
4	ICV	1	Water	NA	NA	20.3	6.98	03/23/2021	08:15
5	CCV1	1	Water	NA	NA	20.3	2.01	03/23/2021	08:16
6	M1764-06	1	Water	NA	NA	21.2	6.61	03/23/2021	08:18
7	M1764-06DUP	1	Water	NA	NA	21.3	6.62	03/23/2021	08:19
8	M1764-10	1	Water	NA	NA	21.0	6.87	03/23/2021	08:20
9	M1764-11	1	Water	NA	NA	21.1	6.81	03/23/2021	08:22
10	M1764-12	1	Water	NA	NA	21.1	6.75	03/23/2021	08:24
11	CCV2	1	Water	NA	NA	21.1	12.02	03/23/2021	08:25

WORKLIST(Hardcopy Internal Chain)

113628 113628

WorkList Name : PH M1764 WATER

WorkList ID : 147367

Department : Wet-Chemistry

Date : 03-23-2021 08:03:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/22/2021	Water	M1764-06	pH	Cool 4 deg C	PSEG03	J21	COMP-OILY-WATER	03/22/2021	9040C
03/22/2021	Water	M1764-10	pH	Cool 4 deg C	PSEG03	J21	BUR-1016	03/22/2021	9040C
03/22/2021	Water	M1764-11	pH	Cool 4 deg C	PSEG03	J21	BUR-1015	03/22/2021	9040C
03/22/2021	Water	M1764-12	pH	Cool 4 deg C	PSEG03	J21	BUR-1014	03/22/2021	9040C

Date/Time 03/23/21 8:05
 Received by: JP (Wet Chem)
 Relinquished by: KS (Wet Chem)

Date/Time 03/22/21 10:00
 Received by: KS (Wet Chem)
 Relinquished by: JP (Wet Chem)