


LB113632

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.3547e-004

C= 3.5450e-002

Rho= 0.9992747

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.058	0.058	94	0.000	94					-10
0.2	0.200	0.180	-0.020	612	0.0 %	612					-1
2.5	2.500	2.485	-0.015	10402	0.0 %	10402					2
5.0	5.000	5.113	0.113	21562	0.0 %	21562					-4
7.5	7.500	7.217	-0.283	30500	0.0 %	30500					1
10.0	10.000	10.148	0.148	42946	0.0 %	42946					

LB113632 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Mar 2021 11:31:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	94	-	0.0000	7470A	PPB	23 Mar 2021 11:39:32	S	Std
0.2 - 1	50.2	612	-	0.2000	7470A	PPB	23 Mar 2021 11:41:48	S	Std
2.5 - 1	52.5	10402	-	2.5000	7470A	PPB	23 Mar 2021 11:44:05	S	Std
5.0 - 1	55	21562	-	5.0000	7470A	PPB	23 Mar 2021 11:46:21	S	Std
7.5 - 1	57.5	30500	-	7.5000	7470A	PPB	23 Mar 2021 11:48:38	S	Std
10.0 - 1	51.8	42946	-	10.0000	7470A	PPB	23 Mar 2021 11:50:53	S	Std
ICV47 - 1	ICV47	15555	3.6982	-	7470A	PPB	23 Mar 2021 11:54:07	U	SMPL
ICB47 - 1	ICB47	18	0.0397	-	7470A	PPB	23 Mar 2021 11:56:22	U	SMPL
CCV59 - 1	CCV59	20691	4.9076	-	7470A	PPB	23 Mar 2021 11:58:37	U	SMPL
CCB59 - 1	CCB59	19	0.0399	-	7470A	PPB	23 Mar 2021 12:00:51	U	SMPL
CRA - 1	CRA	839	0.2330	-	7470A	PPB	23 Mar 2021 12:03:06	U	SMPL
HighStd - 1	HighStd	42414	10.0227	-	7470A	PPB	23 Mar 2021 12:05:21	U	SMPL
ChkStd - 1	ChkStd	28588	6.7671	-	7470A	PPB	23 Mar 2021 12:07:36	U	SMPL
PB135278BL - 1	PBW	55	0.0484	-	7470A	PPB	23 Mar 2021 12:12:23	U	SMPL
PB135278BS - 1	LCSW	15116	3.5948	-	7470A	PPB	23 Mar 2021 12:14:38	U	SMPL
M1713-06 - 1	TP-6A	279	0.1011	-	7470A	PPB	23 Mar 2021 12:16:54	U	SMPL
M1713-07 - 1	TP-7A	289	0.1035	-	7470A	PPB	23 Mar 2021 12:19:11	U	SMPL
M1724-02 - 1	26207	155	0.0719	-	7470A	PPB	23 Mar 2021 12:21:28	U	SMPL
M1724-02DUP - 1	26207DUP	151	0.0710	-	7470A	PPB	23 Mar 2021 12:23:44	U	SMPL
M1724-02MS - 1	26207MS	15988	3.8002	-	7470A	PPB	23 Mar 2021 12:26:00	U	SMPL
CCV60 - 1	CCV60	21182	5.0232	-	7470A	PPB	23 Mar 2021 12:28:15	U	SMPL
CCB60 - 1	CCB60	4	0.0364	-	7470A	PPB	23 Mar 2021 12:30:30	U	SMPL
M1724-02MSD - 1	26207MSD	15888	3.7766	-	7470A	PPB	23 Mar 2021 12:32:45	U	SMPL
PB135279BL - 1	PBW	3	0.0362	-	7470A	PPB	23 Mar 2021 12:35:01	U	SMPL
PB135279BS - 1	LCSW	14697	3.4962	-	7470A	PPB	23 Mar 2021 12:37:16	U	SMPL
M1750-02 - 1	TRE-1015	348	0.1174	-	7470A	PPB	23 Mar 2021 12:39:31	U	SMPL
M1750-02DUP - 1	TRE-1015DUP	355	0.1190	-	7470A	PPB	23 Mar 2021 12:41:46	U	SMPL
M1750-02MS - 1	TRE-1015MS	14791	3.5183	-	7470A	PPB	23 Mar 2021 12:47:05	U	SMPL
M1750-02MSD - 1	TRE-1015MSD	14977	3.5621	-	7470A	PPB	23 Mar 2021 12:49:21	U	SMPL
PB135280BL - 1	PBW	-28	0.0289	-	7470A	PPB	23 Mar 2021 12:51:37	U	SMPL
PB135280BS - 1	LCSW	15570	3.7017	-	7470A	PPB	23 Mar 2021 12:53:54	U	SMPL
M1752-13 - 1	30816	1254	0.3307	-	7470A	PPB	23 Mar 2021 12:56:09	U	SMPL
CCV61 - 1	CCV61	21513	5.1011	-	7470A	PPB	23 Mar 2021 12:58:25	U	SMPL
CCB61 - 1	CCB61	-66	0.0199	-	7470A	PPB	23 Mar 2021 13:00:42	U	SMPL
M1752-13DUP - 1	30816DUP	1096	0.2935	-	7470A	PPB	23 Mar 2021 13:02:57	U	SMPL
M1752-13MS - 1	30816MS	14818	3.5247	-	7470A	PPB	23 Mar 2021 13:07:48	U	SMPL
M1752-13MSD - 1	30816MSD	14348	3.4140	-	7470A	PPB	23 Mar 2021 13:10:03	U	SMPL
M1754-01 - 1	N48987	285	0.1026	-	7470A	PPB	23 Mar 2021 13:12:18	U	SMPL
PB135278TB - 1	PB135272TB	250	0.0943	-	7470A	PPB	23 Mar 2021 13:14:33	U	SMPL
CCV62 - 1	CCV62	20554	4.8753	-	7470A	PPB	23 Mar 2021 13:19:05	U	SMPL
CCB62 - 1	CCB62	-28	0.0289	-	7470A	PPB	23 Mar 2021 13:21:22	U	SMPL
M1724-02LX5 - 1		20	0.0402	-	7470A	PPB	23 Mar 2021 13:23:38	U	SMPL
M1724-02A - 1		15427	3.6681	-	7470A	PPB	23 Mar 2021 13:25:54	U	SMPL
M1750-02LX5 - 1		41	0.0451	-	7470A	PPB	23 Mar 2021 13:28:09	U	SMPL
M1750-02A - 1		15359	3.6521	-	7470A	PPB	23 Mar 2021 13:30:25	U	SMPL
M1752-13LX5 - 1		359	0.1200	-	7470A	PPB	23 Mar 2021 13:32:40	U	SMPL
M1752-13A - 1		18469	4.3844	-	7470A	PPB	23 Mar 2021 13:34:55	U	SMPL
PB135279TB - 1	PB135264TB	207	0.0842	-	7470A	PPB	23 Mar 2021 13:49:36	U	SMPL
CCV63 - 1	CCV63	19974	4.7388	-	7470A	PPB	23 Mar 2021 13:52:35	U	SMPL
CCB63 - 1	CCB63	-37	0.0267	-	7470A	PPB	23 Mar 2021 13:54:50	U	SMPL