

LB113638


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37798



A= 0.0000e+000

B= 2.6697e-004

C= -2.0491e-002

Rho= 0.9996911

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.012	0.012	123	0.000	123					
0.2	0.200	0.163	-0.037	688	0.0 %	688					-19
2.5	2.500	2.528	0.028	9546	0.0 %	9546					1
5.0	5.000	5.105	0.105	19199	0.0 %	19199					2
7.5	7.500	7.321	-0.179	27498	0.0 %	27498					-2
10.0	10.000	10.071	0.071	37798	0.0 %	37798					1

LB113638 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 23 Mar 2021 14:30:39				
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	50	123	-	0.00007470A	PPB	23 Mar 2021 14:32:03	S	Std
0.2 - 1	50.2	688	-	0.20007470A	PPB	23 Mar 2021 14:34:20	S	Std
2.5 - 1	52.5	9546	-	2.50007470A	PPB	23 Mar 2021 14:36:37	S	Std
5.0 - 1	55	19199	-	5.00007470A	PPB	23 Mar 2021 14:38:53	S	Std
7.5 - 1	57.5	27498	-	7.50007470A	PPB	23 Mar 2021 14:41:09	S	Std
10.0 - 1	57.6	37798	-	10.00007470A	PPB	23 Mar 2021 14:43:25	S	Std
ICV49 - 1	ICV49	15113	4.0143	-7470A	PPB	23 Mar 2021 14:46:10	U	SMPL
ICB49 - 1	ICB49	-29	-0.0282	-7470A	PPB	23 Mar 2021 14:48:25	U	SMPL
CCV67 - 1	CCV67	19272	5.1246	-7470A	PPB	23 Mar 2021 14:50:40	U	SMPL
CCB67 - 1	CCB67	-30	-0.0285	-7470A	PPB	23 Mar 2021 14:52:54	U	SMPL
CRA - 1	CRA	648	0.1525	-7470A	PPB	23 Mar 2021 14:55:09	U	SMPL
HighStd - 1	HighStd	37516	9.9953	-7470A	PPB	23 Mar 2021 14:57:24	U	SMPL
ChkStd - 1	ChkStd	26695	7.1064	-7470A	PPB	23 Mar 2021 14:59:39	U	SMPL
PB135309BL - 1	PBW	-69	-0.0389	-7470A	PPB	23 Mar 2021 15:01:55	U	SMPL
PB135309BS - 1	LCSW	13917	3.6950	-7470A	PPB	23 Mar 2021 15:04:10	U	SMPL
M1761-01 - 1	TREMONT-STATION-DRUM-5-PPE	-4	-0.0216	-7470A	PPB	23 Mar 2021 15:06:25	U	SMPL
M1761-02 - 1	TREMONT-STATION-DRUM-6-POLY	7	-0.0186	-7470A	PPB	23 Mar 2021 15:08:41	U	SMPL
M1764-01 - 1	BUR-1009	29	-0.0127	-7470A	PPB	23 Mar 2021 15:10:58	U	SMPL
M1764-01DUP - 1	BUR-1009DUP	41	-0.0095	-7470A	PPB	23 Mar 2021 15:13:15	U	SMPL
M1764-01MS - 1	BUR-1009MS	14267	3.7884	-7470A	PPB	23 Mar 2021 15:15:30	U	SMPL
CCV68 - 1	CCV68	19272	5.1246	-7470A	PPB	23 Mar 2021 15:17:45	U	SMPL
CCB68 - 1	CCB68	-29	-0.0282	-7470A	PPB	23 Mar 2021 15:20:01	U	SMPL
M1764-01MSD - 1	BUR-1009MSD	13655	3.6250	-7470A	PPB	23 Mar 2021 15:22:16	U	SMPL
PB135309TB - 1	PB135282TB	-22	-0.0264	-7470A	PPB	23 Mar 2021 15:24:32	U	SMPL
M1764-01LX5 - 1		27	-0.0133	-7470A	PPB	23 Mar 2021 15:26:47	U	SMPL
M1764-01A - 1		14827	3.9379	-7470A	PPB	23 Mar 2021 15:29:02	U	SMPL
CCV69 - 1	CCV69	19423	5.1649	-7470A	PPB	23 Mar 2021 15:31:18	U	SMPL
CCB69 - 1	CCB69	-29	-0.0282	-7470A	PPB	23 Mar 2021 15:33:35	U	SMPL