



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113647

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/23/2021	16:25
2	CAL2	1	Water	NA	NA	20.2	7.00	03/23/2021	16:26
3	CAL3	1	Water	NA	NA	20.2	10.03	03/23/2021	16:27
4	ICV	1	Water	NA	NA	20.3	6.99	03/23/2021	16:30
5	CCV1	1	Water	NA	NA	20.3	2.01	03/23/2021	16:32
6	M1774-01	1	Water	NA	NA	20.4	4.89	03/23/2021	16:33
7	M1774-01DUP	1	Water	NA	NA	20.5	4.90	03/23/2021	16:34
8	M1774-02	1	Water	NA	NA	20.3	6.65	03/23/2021	16:35
9	M1774-03	1	Water	NA	NA	20.4	6.33	03/23/2021	16:36
10	M1774-04	1	Water	NA	NA	20.3	6.90	03/23/2021	16:37
11	M1776-05	1	Water	NA	NA	20.4	7.66	03/23/2021	16:39
12	M1777-04	1	Water	NA	NA	20.1	6.91	03/23/2021	16:40
13	M1777-06	1	Water	NA	NA	20.5	7.60	03/23/2021	16:42
14	M1777-07	1	Water	NA	NA	20.3	7.50	03/23/2021	16:43
15	CCV2	1	Water	NA	NA	20.4	12.02	03/23/2021	16:45

WORKLIST(Hardcopy Internal Chain)

VB 113647

WorkList Name : PH M1774 WorkList ID : 147400 Department : Wet-Chemistry Date : 03-23-2021 15:51:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/23/2021	Water	M1774-01	pH	Cool 4 deg C	PSEG03	J42	41736	03/23/2021	9040C
03/23/2021	Water	M1774-02	pH	Cool 4 deg C	PSEG03	J42	30806	03/23/2021	9040C
03/23/2021	Water	M1774-03	pH	Cool 4 deg C	PSEG03	J42	30807	03/23/2021	9040C
03/23/2021	Water	M1774-04	pH	Cool 4 deg C	PSEG03	J42	30808	03/23/2021	9040C
03/23/2021	Water	M1776-05	pH	Cool 4 deg C	PSEG03	J43	107	03/23/2021	9040C
03/23/2021	Water	M1777-04	pH	Cool 4 deg C	PSEG03	J42	245	03/23/2021	9040C
03/23/2021	Water	M1777-06	pH	Cool 4 deg C	PSEG03	J42	238	03/23/2021	9040C
03/23/2021	Water	M1777-07	pH	Cool 4 deg C	PSEG03	J42	243	03/23/2021	9040C

Date/Time 03/23/21 16:15
 Received by: JO (wet-chem)
 Relinquished by: KS (wet-chem)

Date/Time 03/23/21 17:40
 Received by: KS (wet-chem)
 Relinquished by: JO (wet-chem)