

LB113649


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35292



A= 0.0000e+000

B= 2.8241e-004

C= -1.9490e-002

Rho= 0.9998928

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.006	-0.006	46	0.000	46					-10
0.05	0.050	0.040	-0.010	210	0.0 %	210	0				-13
0.2	0.200	0.174	-0.026	684	0.0 %	684					-1
2.5	2.500	2.477	-0.023	8840	0.0 %	8840					3
5.0	5.000	5.128	0.128	18228	0.0 %	18228					0
7.5	7.500	7.491	-0.009	26593	0.0 %	26593					-1
10.0	10.000	9.947	-0.053	35292	0.0 %	35292					

LB113649 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 24 Mar 2021 11:19:10

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	46	-	0.0000	245.1	PPB	24 Mar 2021 11:27:04	S	Std
0.05 - 1	50.05	210	-	0.0500	245.1	PPB	24 Mar 2021 11:29:20	S	Std
0.2 - 1	50.2	684	-	0.2000	245.1	PPB	24 Mar 2021 11:31:38	S	Std
2.5 - 1	52.5	8840	-	2.5000	245.1	PPB	24 Mar 2021 11:33:54	S	Std
5.0 - 1	55	18228	-	5.0000	245.1	PPB	24 Mar 2021 11:36:10	S	Std
7.5 - 1	57.5	26593	-	7.5000	245.1	PPB	24 Mar 2021 11:40:59	S	Std
10.0 - 1	510	35292	-	10.0000	245.1	PPB	24 Mar 2021 11:45:50	S	Std
ICV50 - 1	ICV50	14067	3.9531	-245.1		PPB	24 Mar 2021 11:50:36	U	SMPL
ICB50 - 1	ICB50	-15	-0.0237	-245.1		PPB	24 Mar 2021 11:52:50	U	SMPL
CCV70 - 1	CCV70	18930	5.3265	-245.1		PPB	24 Mar 2021 11:55:05	U	SMPL
CCB70 - 1	CCB70	-43	-0.0316	-245.1		PPB	24 Mar 2021 11:57:20	U	SMPL
CRA - 1	CRA	224	0.0438	-245.1		PPB	24 Mar 2021 12:02:06	U	SMPL
HighStd - 1	HighStd	36627	10.3242	-245.1		PPB	24 Mar 2021 12:04:21	U	SMPL
ChkStd - 1	ChkStd	25876	7.2881	-245.1		PPB	24 Mar 2021 12:06:36	U	SMPL
PB135302BL - 1	PBW	-58	-0.0359	-245.1		PPB	24 Mar 2021 12:08:52	U	SMPL
PB135302BS - 1	LCSW	13983	3.9294	-245.1		PPB	24 Mar 2021 12:11:08	U	SMPL
M1704-02 - 1	SPRING-2021	271	0.0570	-245.1		PPB	24 Mar 2021 12:13:24	U	SMPL
M1704-02DUP - 1	SPRING-2021DUP	268	0.0562	-245.1		PPB	24 Mar 2021 12:15:41	U	SMPL
M1704-02MS - 1	SPRING-2021MS	13096	3.6789	-245.1		PPB	24 Mar 2021 12:20:31	U	SMPL
M1704-02MSD - 1	SPRING-2021MSD	14283	4.0141	-245.1		PPB	24 Mar 2021 12:22:47	U	SMPL
M1710-01 - 1	EFF-WW	20	-0.0138	-245.1		PPB	24 Mar 2021 12:25:03	U	SMPL
CCV71 - 1	CCV71	18444	5.1892	-245.1		PPB	24 Mar 2021 12:27:18	U	SMPL
CCB71 - 1	CCB71	-57	-0.0356	-245.1		PPB	24 Mar 2021 12:29:33	U	SMPL
M1739-06 - 1	HP-627	510	0.1245	-245.1		PPB	24 Mar 2021 12:31:48	U	SMPL
M1740-01 - 1	TI-604-EB	52	-0.0048	-245.1		PPB	24 Mar 2021 12:34:03	U	SMPL
M1740-03 - 1	TI-604	202	0.0376	-245.1		PPB	24 Mar 2021 12:36:18	U	SMPL
M1740-04 - 1	TI-604-X	181	0.0316	-245.1		PPB	24 Mar 2021 12:38:33	U	SMPL
M1764-06 - 1	COMP-OILY-WATER	333	0.0746	-245.1		PPB	24 Mar 2021 12:40:49	U	SMPL
M1704-02LX5 - 1		83	0.0040	-245.1		PPB	24 Mar 2021 12:43:06	U	SMPL
M1704-02A - 1		13567	3.8119	-245.1		PPB	24 Mar 2021 12:45:22	U	SMPL
CCV72 - 1	CCV72	18395	5.1754	-245.1		PPB	24 Mar 2021 12:47:39	U	SMPL
CCB72 - 1	CCB72	-23	-0.0260	-245.1		PPB	24 Mar 2021 12:49:56	U	SMPL