

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AlexanderParameter: Phosphorus, TotalSUPERVISOR REVIEW BY: apatelRun Number: LB113650BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP90157
calibration std. phosphate 0.5 ppm	WP90156
calibration std. phosphate 0.3 ppm	WP90155
calibration std. phosphate 0.1 ppm	WP90154
calibration std. phosphate 0.05 ppm	WP90153
calibration std. 0 ppm	WP90152
phosphate CCV std.	WP90158
5N sulfuric acid	WP88023
Combined reagent	WP90163
Phenolphthalein indicator	WP88839
Sodium hydroxide, 1N	WP89699
Phosphate ICV-LCS Std	WP90159
Phosphate LOD-MDL Std 0.025ppm	WP90161

Intercept: 0.0018Slope: 0.631Regression: 0.999978

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.001	-0.001		03/24/2021	12:36
2	CAL2	0.05	1	50	50	0.035	0.053	-6	03/24/2021	12:37
3	CAL3	0.10	1	50	50	0.066	0.102	-2	03/24/2021	12:37
4	CAL4	0.30	1	50	50	0.190	0.298	0.7	03/24/2021	12:38
5	CAL5	0.50	1	50	50	0.315	0.496	0.8	03/24/2021	12:38
6	CAL6	1.00	1	50	50	0.634	1.002	-0.2	03/24/2021	12:39



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Alexander

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB113650

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.318	0.501	03/24/2021	12:40
2	ICB		1	50	50	0.001	-0.001	03/24/2021	12:40
3	CCV1	0.50	1	50	50	0.313	0.493	03/24/2021	12:41
4	CCB1		1	50	50	0.002	0.000	03/24/2021	12:42
5	PB135332BL		1	1.00	50	0.002	0.000	03/24/2021	12:42
6	PB135332BS	0.50	1	1.00	50	0.320	0.504	03/24/2021	12:43
7	M1458-03		1	1.00	50	0.021	0.030	03/24/2021	12:44
8	CCV2	0.50	1	50	50	0.313	0.493	03/24/2021	12:44
9	CCB2		1	50	50	0.002	0.000	03/24/2021	12:44