



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB113651

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/24/2021	13:00
2	CAL2	1	Water	NA	NA	20.3	6.99	03/24/2021	13:01
3	CAL3	1	Water	NA	NA	20.3	10.04	03/24/2021	13:02
4	ICV	1	Water	NA	NA	20.3	7.01	03/24/2021	13:05
5	CCV1	1	Water	NA	NA	20.1	2.02	03/24/2021	13:06
6	M1796-02	1	Water	NA	NA	20.4	11.87	03/24/2021	13:09
7	M1796-02DUP	1	Water	NA	NA	20.5	11.88	03/24/2021	13:10
8	CCV2	1	Water	NA	NA	20.1	12.02	03/24/2021	13:11

2B113651

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M1596 V WorkList ID : 147430 Department : Wet-Chemistry Date : 03-24-2021 12:46:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/23/2021	Water	M1796-02	Corrosivity	Cool 4 deg C	ARCA05	L12	PB-2021-WWW-000-06	03/23/2021	9040C

Date/Time 03/24/21 12:50
Received by: JP (wet Chem)
Relinquished by: JP (wet Chem)

Date/Time 03/24/21 14:00
Received by: JP (wet Chem)
Relinquished by: JP (wet Chem)