


LB 113 653

7471B

INSTRUMENT ID: CV1

 μ Abs.:

59209



A= 0.0000e+000

B= 1.7043e-004

C= 2.4245e-002

Rho= 0.9997525

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.034	0.034	55	0.000	55					
0.2	0.200	0.206	0.006	1064	0.0 %	1064					
2.5	2.500	2.507	0.007	14570	0.0 %	14570					
5.0	5.000	4.999	-0.001	29192	0.0 %	29192					
7.5	7.500	7.339	-0.161	42919	0.0 %	42919					
10.0	10.000	10.115	0.115	59209	0.0 %	59209					

3
-2
1

LB113653 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Mar 2021 14:46:37

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	55	-	0.00007471B	PPB		24 Mar 2021 15:05:27	S	Std
0.2 - 1	50.2	1064	-	0.20007471B	PPB		24 Mar 2021 15:07:43	S	Std
2.5 - 1	52.5	14570	-	2.50007471B	PPB		24 Mar 2021 15:10:00	S	Std
5.0 - 1	55	29192	-	5.00007471B	PPB		24 Mar 2021 15:12:17	S	Std
7.5 - 1	57.5	42919	-	7.50007471B	PPB		24 Mar 2021 15:14:33	S	Std
10.0 - 1	510	59209	-	10.00007471B	PPB		24 Mar 2021 15:22:00	S	Std
ICV51 - 1	ICV51	21857	3.7493	-7471B	PPB		24 Mar 2021 15:24:46	U	SMPL
ICB51 - 1	ICB51	-53	0.0152	-7471B	PPB		24 Mar 2021 15:27:00	U	SMPL
CCV73 - 1	CCV73	29401	5.0350	-7471B	PPB		24 Mar 2021 15:29:15	U	SMPL
CCB73 - 1	CCB73	-76	0.0113	-7471B	PPB		24 Mar 2021 15:31:30	U	SMPL
CRA - 1	CRA	1057	0.2044	-7471B	PPB		24 Mar 2021 15:33:45	U	SMPL
HighStd - 1	HighStd	60752	10.3781	-7471B	PPB		24 Mar 2021 15:36:00	U	SMPL
ChkStd - 1	ChkStd	42086	7.1969	-7471B	PPB		24 Mar 2021 15:38:15	U	SMPL
PB135303BL - 1	PBS	-148	-0.0010	-7471B	PPB		24 Mar 2021 15:40:33	U	SMPL
PB135303BS - 1	LCSS	21253	3.6464	-7471B	PPB		24 Mar 2021 15:42:49	U	SMPL
M1730-01 - 1	TP-8A	147389	25.1435	-7471B	PPB		24 Mar 2021 15:45:05	U	SMPL
M1765-02 - 1	RBR-07	8868	1.5356	-7471B	PPB		24 Mar 2021 15:47:21	U	SMPL
M1765-04 - 1	RBR-54	9565	1.6544	-7471B	PPB		24 Mar 2021 15:49:40	U	SMPL
M1766-01 - 1	OK-03-032221	1491	0.2784	-7471B	PPB		24 Mar 2021 15:51:57	U	SMPL
CCV74 - 1	CCV74	29440	5.0417	-7471B	PPB		24 Mar 2021 15:54:12	U	SMPL
CCB74 - 1	CCB74	-54	0.0150	-7471B	PPB		24 Mar 2021 15:56:27	U	SMPL
M1768-01 - 1	EO-02-032221	4213	0.7423	-7471B	PPB		24 Mar 2021 16:03:35	U	SMPL
M1768-01DUP - 1	EO-02-032221DUP	3977	0.7020	-7471B	PPB		24 Mar 2021 16:05:50	U	SMPL
M1768-01MS - 1	EO-02-032221MS	25181	4.3158	-7471B	PPB		24 Mar 2021 16:08:06	U	SMPL
M1768-01MSD - 1	EO-02-032221MSD	24994	4.2839	-7471B	PPB		24 Mar 2021 16:13:00	U	SMPL
M1771-01 - 1	COMP-SOIL-323	704	0.1442	-7471B	PPB		24 Mar 2021 16:15:15	U	SMPL
M1773-01 - 1	NB-08-032321	1923	0.3520	-7471B	PPB		24 Mar 2021 16:17:31	U	SMPL
M1786-01 - 1	SP-1	1107	0.2129	-7471B	PPB		24 Mar 2021 16:19:47	U	SMPL
M1768-01LX5 - 1		860	0.1708	-7471B	PPB		24 Mar 2021 16:22:03	U	SMPL
M1768-01A - 1		28958	4.9595	-7471B	PPB		24 Mar 2021 16:24:19	U	SMPL
M1730-01DLX5 - 1		30237	5.1775	-7471B	PPB		24 Mar 2021 16:26:36	U	SMPL
CCV75 - 1	CCV75	29546	5.0597	-7471B	PPB		24 Mar 2021 16:28:52	U	SMPL
CCB75 - 1	CCB75	-38	0.0178	-7471B	PPB		24 Mar 2021 16:31:07	U	SMPL