

7470A


LB113664

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39731



A= 0.0000e+000

B= 2.5032e-004

C= -4.1096e-003

Rho= 0.9998059

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.019	0.019	91	0.000	91					-8
0.2	0.200	0.185	-0.015	754	0.0 %	754					-4
2.5	2.500	2.404	-0.096	9620	0.0 %	9620					3
5.0	5.000	5.135	0.135	20530	0.0 %	20530					0
7.5	7.500	7.516	0.016	30042	0.0 %	30042					-1
10.0	10.000	9.942	-0.058	39731	0.0 %	39731					

LB113664 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 25 Mar 2021 10:02:42					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	91	-	0.00007470A	PPB		25 Mar 2021 10:03:03	S	Std
0.2 - 1	50.2	754	-	0.20007470A	PPB		25 Mar 2021 10:05:20	S	Std
2.5 - 1	52.5	9620	-	2.50007470A	PPB		25 Mar 2021 10:07:37	S	Std
5.0 - 1	55	20530	-	5.00007470A	PPB		25 Mar 2021 10:09:53	S	Std
7.5 - 1	57.5	30042	-	7.50007470A	PPB		25 Mar 2021 10:12:09	S	Std
10.0 - 1	510	39731	-	10.00007470A	PPB		25 Mar 2021 10:17:04	S	Std
ICV52 - 1	ICV52	15713	3.9292	-7470A	PPB		25 Mar 2021 10:19:47	U	SMPL
ICB52 - 1	ICB52	-31	-0.0119	-7470A	PPB		25 Mar 2021 10:22:02	U	SMPL
CCV76 - 1	CCV76	20272	5.0705	-7470A	PPB		25 Mar 2021 10:24:17	U	SMPL
CCB76 - 1	CCB76	-28	-0.0111	-7470A	PPB		25 Mar 2021 10:26:32	U	SMPL
CRA - 1	CRA	742	0.1816	-7470A	PPB		25 Mar 2021 10:28:46	U	SMPL
HighStd - 1	HighStd	39989	10.0061	-7470A	PPB		25 Mar 2021 10:31:01	U	SMPL
ChkStd - 1	ChkStd	28450	7.1176	-7470A	PPB		25 Mar 2021 10:33:16	U	SMPL
PB135335BL - 1	PBW	-73	-0.0224	-7470A	PPB		25 Mar 2021 10:35:32	U	SMPL
PB135335BS - 1	LCSW	13869	3.4676	-7470A	PPB		25 Mar 2021 10:37:47	U	SMPL
M1730-03 - 1	TP-8A	-31	-0.0119	-7470A	PPB		25 Mar 2021 10:40:03	U	SMPL
M1730-03DUP - 1	TP-8ADUP	58	0.0104	-7470A	PPB		25 Mar 2021 10:42:20	U	SMPL
M1730-03MS - 1	TP-8AMS	14852	3.7137	-7470A	PPB		25 Mar 2021 10:44:37	U	SMPL
M1730-03MSD - 1	TP-8AMSD	15261	3.8161	-7470A	PPB		25 Mar 2021 10:46:53	U	SMPL
M1771-05 - 1	COMP-SOIL-323	-37	-0.0134	-7470A	PPB		25 Mar 2021 10:49:09	U	SMPL
CCV77 - 1	CCV77	20539	5.1373	-7470A	PPB		25 Mar 2021 10:51:24	U	SMPL
CCB77 - 1	CCB77	-36	-0.0131	-7470A	PPB		25 Mar 2021 10:53:40	U	SMPL
M1772-01 - 1	DRUM-1-2-COMP	21	0.0011	-7470A	PPB		25 Mar 2021 10:55:55	U	SMPL
M1772-02 - 1	550026646	78	0.0154	-7470A	PPB		25 Mar 2021 10:58:10	U	SMPL
M1776-01 - 1	110	29	0.0031	-7470A	PPB		25 Mar 2021 11:00:26	U	SMPL
M1776-03 - 1	106	70	0.0134	-7470A	PPB		25 Mar 2021 11:02:42	U	SMPL
M1777-01 - 1	244	54	0.0094	-7470A	PPB		25 Mar 2021 11:04:57	U	SMPL
M1777-02 - 1	248-250	33	0.0042	-7470A	PPB		25 Mar 2021 11:07:13	U	SMPL
PB135336BL - 1	PBW	31	0.0037	-7470A	PPB		25 Mar 2021 11:09:29	U	SMPL
PB135336BS - 1	LCSW	13547	3.3870	-7470A	PPB		25 Mar 2021 11:11:46	U	SMPL
M1776-05 - 1	107	11	-0.0014	-7470A	PPB		25 Mar 2021 11:14:02	U	SMPL
M1776-05DUP - 1	107DUP	41	0.0062	-7470A	PPB		25 Mar 2021 11:16:18	U	SMPL
CCV78 - 1	CCV78	18566	4.6434	-7470A	PPB		25 Mar 2021 11:18:33	U	SMPL
CCB78 - 1	CCB78	-24	-0.0101	-7470A	PPB		25 Mar 2021 11:20:48	U	SMPL
M1776-05MS - 1	107MS	13386	3.3467	-7470A	PPB		25 Mar 2021 11:23:04	U	SMPL
M1776-05MSD - 1	107MSD	13815	3.4541	-7470A	PPB		25 Mar 2021 11:25:19	U	SMPL
M1777-04 - 1	245	-21	-0.0094	-7470A	PPB		25 Mar 2021 11:27:35	U	SMPL
M1777-06 - 1	238	33	0.0042	-7470A	PPB		25 Mar 2021 11:29:50	U	SMPL
M1777-07 - 1	243	172	0.0389	-7470A	PPB		25 Mar 2021 11:32:06	U	SMPL
PB135337BL - 1	PBW	-11	-0.0069	-7470A	PPB		25 Mar 2021 11:34:22	U	SMPL
PB135337BS - 1	LCSW	14747	3.6874	-7470A	PPB		25 Mar 2021 11:50:11	U	SMPL
M1770-02 - 1	MW-106I	46	0.0074	-7470A	PPB		25 Mar 2021 11:52:27	U	SMPL
M1770-03 - 1	MW-106S	52	0.0089	-7470A	PPB		25 Mar 2021 11:54:44	U	SMPL
M1770-03DUP - 1	MW-106SDUP	50	0.0084	-7470A	PPB		25 Mar 2021 11:56:59	U	SMPL
CCV79 - 1	CCV79	18988	4.7490	-7470A	PPB		25 Mar 2021 11:59:14	U	SMPL
CCB79 - 1	CCB79	0	-0.0041	-7470A	PPB		25 Mar 2021 12:01:30	U	SMPL
M1770-04 - 1	M1770-03MS	13565	3.3915	-7470A	PPB		25 Mar 2021 12:03:45	U	SMPL
M1770-05 - 1	M1770-03MSD	15059	3.7655	-7470A	PPB		25 Mar 2021 12:06:00	U	SMPL
M1770-06 - 1	MW-105D	175	0.0397	-7470A	PPB		25 Mar 2021 12:08:15	U	SMPL
M1770-07 - 1	MW-105I	30	0.0034	-7470A	PPB		25 Mar 2021 12:10:31	U	SMPL
M1770-08 - 1	MW-105S	10	-0.0016	-7470A	PPB		25 Mar 2021 12:12:46	U	SMPL
M1770-10 - 1	MW-106D	59	0.0107	-7470A	PPB		25 Mar 2021 12:15:03	U	SMPL
M1770-11 - 1	MW-104D	277	0.0652	-7470A	PPB		25 Mar 2021 12:17:19	U	SMPL
M1770-12 - 1	MW-104I	8	-0.0021	-7470A	PPB		25 Mar 2021 12:19:36	U	SMPL
M1770-13 - 1	MW-103D	60	0.0109	-7470A	PPB		25 Mar 2021 12:21:53	U	SMPL
M1770-14 - 1	MW-103I	116	0.0249	-7470A	PPB		25 Mar 2021 12:24:09	U	SMPL
CCV80 - 1	CCV80	18829	4.7092	-7470A	PPB		25 Mar 2021 12:26:24	U	SMPL
CCB80 - 1	CCB80	-5	-0.0054	-7470A	PPB		25 Mar 2021 12:28:40	U	SMPL
M1770-15 - 1	MW-101S	464	0.1120	-7470A	PPB		25 Mar 2021 12:30:55	U	SMPL
M1770-16 - 1	MW-101I	70	0.0134	-7470A	PPB		25 Mar 2021 12:33:11	U	SMPL
M1770-17 - 1	MW-101D	97	0.0202	-7470A	PPB		25 Mar 2021 12:35:26	U	SMPL
M1770-18 - 1	FB-03232021	47	0.0077	-7470A	PPB		25 Mar 2021 12:37:42	U	SMPL
M1774-01 - 1	41736	1170	0.2888	-7470A	PPB		25 Mar 2021 12:39:57	U	SMPL
M1774-02 - 1	30806	1833	0.4547	-7470A	PPB		25 Mar 2021 12:42:14	U	SMPL
M1796-02 - 1	PB-2021-WW-000-06	67	0.0127	-7470A	PPB		25 Mar 2021 12:44:30	U	SMPL
PB135335TB - 1	PB135316TB	-7	-0.0059	-7470A	PPB		25 Mar 2021 12:46:46	U	SMPL
PB135336TB - 1	PB135319TB	27	0.0026	-7470A	PPB		25 Mar 2021 12:49:03	U	SMPL
CCV81 - 1	CCV81	19126	4.7836	-7470A	PPB		25 Mar 2021 12:51:18	U	SMPL
CCB81 - 1	CCB81	-9	-0.0064	-7470A	PPB		25 Mar 2021 12:53:34	U	SMPL
M1730-03LX5 - 1		39	0.0057	-7470A	PPB		25 Mar 2021 12:55:49	U	SMPL
M1730-03A - 1		14639	3.6604	-7470A	PPB		25 Mar 2021 12:58:05	U	SMPL
M1776-05LX5 - 1		7	-0.0024	-7470A	PPB		25 Mar 2021 13:00:20	U	SMPL
M1776-05A - 1		14796	3.6997	-7470A	PPB		25 Mar 2021 13:02:36	U	SMPL
M1770-03LX5 - 1		4	-0.0031	-7470A	PPB		25 Mar 2021 13:04:52	U	SMPL
M1770-03A - 1		15739	3.9357	-7470A	PPB		25 Mar 2021 13:07:08	U	SMPL
CCV82 - 1	CCV82	19001	4.7523	-7470A	PPB		25 Mar 2021 13:09:24	U	SMPL
CCB82 - 1	CCB82	0	-0.0041	-7470A	PPB		25 Mar 2021 13:11:40	U	SMPL