

LB113669


7471B

INSTRUMENT ID: CVA

Linear

 μ Abs.:

53456



A= 0.0000e+000

B= 1.8725e-004

C= 3.2131e-003

Rho= 0.9999919

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.019	0.019	83	0.000	83					
0.2	0.200	0.196	-0.004	1029	0.0 %	1029					-2
2.5	2.500	2.499	-0.001	13326	0.0 %	13326					0
5.0	5.000	4.972	-0.028	26536	0.0 %	26536					-1
7.5	7.500	7.502	0.002	40044	0.0 %	40044					0
10.0	10.000	10.013	0.013	53456	0.0 %	53456					0

LB113669 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 25 Mar 2021 15:15:37

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	S0	83	-	0.00007471B	PPB	25 Mar 2021 15:18:13	S	Std
0.2 - 1	52.2	1029	-	0.20007471B	PPB	25 Mar 2021 15:20:30	S	Std
2.5 - 1	52.5	13326	-	2.50007471B	PPB	25 Mar 2021 15:22:47	S	Std
5.0 - 1	55	26536	-	5.00007471B	PPB	25 Mar 2021 15:25:03	S	Std
7.5 - 1	57.5	40044	-	7.50007471B	PPB	25 Mar 2021 15:27:19	S	Std
10.0 - 1	510	53456	-	10.00007471B	PPB	25 Mar 2021 15:29:35	S	Std
ICV53 - 1	ICV53	21373	4.0054	-7471B	PPB	25 Mar 2021 15:32:19	U	SMPL
ICB53 - 1	ICB53	2	0.0036	-7471B	PPB	25 Mar 2021 15:34:34	U	SMPL
CCV83 - 1	CCV83	26870	5.0347	-7471B	PPB	25 Mar 2021 15:36:49	U	SMPL
CCB83 - 1	CCB83	4	0.0040	-7471B	PPB	25 Mar 2021 15:39:04	U	SMPL
CRA - 1	CRA	1012	0.1927	-7471B	PPB	25 Mar 2021 15:41:18	U	SMPL
HighStd - 1	HighStd	53510	10.0231	-7471B	PPB	25 Mar 2021 15:43:33	U	SMPL
ChkStd - 1	ChkStd	38156	7.1481	-7471B	PPB	25 Mar 2021 15:45:48	U	SMPL
PB135359BL - 1	PBS	-76	-0.0110	-7471B	PPB	25 Mar 2021 15:48:05	U	SMPL
PB135359BS - 1	LCSS	20175	3.7811	-7471B	PPB	25 Mar 2021 15:50:21	U	SMPL
M1775-02 - 1	11979	2530	0.4770	-7471B	PPB	25 Mar 2021 15:52:37	U	SMPL
M1792-01 - 1	TP-5	2687	0.5064	-7471B	PPB	25 Mar 2021 15:54:53	U	SMPL
M1792-04 - 1	TP-6	9027	1.6936	-7471B	PPB	25 Mar 2021 15:57:10	U	SMPL
M1793-01 - 1	SU-04-032421	27634	5.1778	-7471B	PPB	25 Mar 2021 15:59:27	U	SMPL
M1793-01DUP - 1	SU-04-032421DUP	27306	5.1164	-7471B	PPB	25 Mar 2021 16:01:42	U	SMPL
CCV84 - 1	CCV84	27148	5.0868	-7471B	PPB	25 Mar 2021 16:03:57	U	SMPL
CCB84 - 1	CCB84	-4	0.0025	-7471B	PPB	25 Mar 2021 16:06:12	U	SMPL
M1793-01MS - 1	SU-04-032421MS	51982	9.7370	-7471B	PPB	25 Mar 2021 16:11:05	U	SMPL
M1793-01MSD - 1	SU-04-032421MSD	50782	9.5123	-7471B	PPB	25 Mar 2021 16:16:05	U	SMPL
M1793-01LX5 - 1		5124	0.9627	-7471B	PPB	25 Mar 2021 16:18:19	U	SMPL
M1793-01A - 1		51470	9.6411	-7471B	PPB	25 Mar 2021 16:20:36	U	SMPL
CCV85 - 1	CCV85	27335	5.1218	-7471B	PPB	25 Mar 2021 16:22:51	U	SMPL
CCB85 - 1	CCB85	-9	0.0015	-7471B	PPB	25 Mar 2021 16:25:08	U	SMPL