



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113670

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	03/25/2021	09:25
2	M1763-01		1	10	100.00	>212.0	>212.0	03/25/2021	09:45
3	M1763-01DUP		1	10	100.00	>212.0	>212.0	03/25/2021	10:10
4	M1764-06		1	12	100.00	>212.0	>212.0	03/25/2021	10:35
5	M1774-01		1	15	100.00	>212.0	>212.0	03/25/2021	10:58
6	M1774-02		1	13	100.00	>212.0	>212.0	03/25/2021	11:20
7	M1774-03		1	14	100.00	>212.0	>212.0	03/25/2021	11:40
8	M1774-04		1	16	100.00	>212.0	>212.0	03/25/2021	12:05
9	M1776-05		1	15	100.00	>212.0	>212.0	03/25/2021	12:32
10	M1777-04		1	16	100.00	>212.0	>212.0	03/25/2021	12:55
11	M1777-06		1	12	100.00	>212.0	>212.0	03/25/2021	13:20
12	M1777-07		1	15	100.00	>212.0	>212.0	03/25/2021	13:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 11 3670

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp32321 WorkList ID : 147402 Department : Wet-Chemistry Date : 03-23-2021 16:33:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/02/2021	Water	M1763-01	Flash Point	Cool 4 deg C	TULL02	J11	EL-509-WATER-SAMPLE	03/19/2021	1010A
04/05/2021	Water	M1764-06	Flash Point	Cool 4 deg C	PSEG03	J21	COMP-OILY-WATER	03/22/2021	1010A
03/30/2021	Water	M1774-01	Flash Point	Cool 4 deg C	PSEG03	J42	41736	03/23/2021	1010A
03/30/2021	Water	M1774-02	Flash Point	Cool 4 deg C	PSEG03	J42	30806	03/23/2021	1010A
03/30/2021	Water	M1774-03	Flash Point	Cool 4 deg C	PSEG03	J42	30807	03/23/2021	1010A
03/30/2021	Water	M1774-04	Flash Point	Cool 4 deg C	PSEG03	J42	30808	03/23/2021	1010A
03/30/2021	Water	M1776-05	Flash Point	Cool 4 deg C	PSEG03	J43	107	03/23/2021	1010A
03/30/2021	Water	M1777-04	Flash Point	Cool 4 deg C	PSEG03	J42	245	03/23/2021	1010A
03/30/2021	Water	M1777-06	Flash Point	Cool 4 deg C	PSEG03	J42	238	03/23/2021	1010A
03/30/2021	Water	M1777-07	Flash Point	Cool 4 deg C	PSEG03	J42	243	03/23/2021	1010A

Date/Time 03/25/21 09:05
Received by: KS (Wetchem)
Relinquished by: JP (Wetchem)

Date/Time 03/25/21 13:50
Received by: JP (Wetchem)
Relinquished by: KS (Wetchem)