



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113685

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/25/2021	17:20
2	CAL2	1	Water	NA	NA	20.2	7.00	03/25/2021	17:21
3	CAL3	1	Water	NA	NA	20.2	10.03	03/25/2021	17:22
4	ICV	1	Water	NA	NA	20.2	6.99	03/25/2021	17:24
5	CCV1	1	Water	NA	NA	20.2	2.01	03/25/2021	17:25
6	M1809-04	1	Water	NA	NA	20.7	7.50	03/25/2021	17:26
7	M1809-05	1	Water	NA	NA	20.6	6.98	03/25/2021	17:28
8	M1809-05DUP	1	Water	NA	NA	20.7	6.99	03/25/2021	17:29
9	CCV2	1	Water	NA	NA	20.3	12.02	03/25/2021	17:30

WORKLIST(Hardcopy Internal Chain)

VB 113685

WorkList Name : PH M1809 W WorkList ID : 147478 Department : Wet-Chemistry Date : 03-25-2021 17:09:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/25/2021	Water	M1809-04	pH ✓	Cool 4 deg C	PSEG03	J61	DRUM-3	03/25/2021	9040C
03/25/2021	Water	M1809-05	pH ✓	Cool 4 deg C	PSEG03	J61	DRUM-4	03/25/2021	9040C

Date/Time 03/25/21 17:12
 Received by: JO (wet chem)
 Relinquished by: AMC wet-chem

Date/Time 03/25/21 18:15
 Received by: AMC wet-chem
 Relinquished by: JO wet chem