

LB113698


7470A

INSTRUMENT ID: CVL

Linear

 μ Abs.:

39591



A= 0.0000e+000

B= 2.4986e-004

C= 3.5039e-002

Rho= 0.9997521

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%OD
0.0	0.000	0.044	0.044	34	0.000	34					
0.2	0.200	0.218	0.018	731	0.0 %	731					9
2.5	2.500	2.362	-0.138	9315	0.0 %	9315					-6
5.0	5.000	5.036	0.036	20016	0.0 %	20016					1
7.5	7.500	7.613	0.113	30329	0.0 %	30329					2
10.0	10.000	9.927	-0.073	39591	0.0 %	39591					-1

LB113693 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 26 Mar 2021 11:39:21					
Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type	
0.0 - 1	50	34	-	0.00007470A	PPB	26 Mar 2021 11:44:48	S	Std	
0.2 - 1	50.2	731	-	0.20007470A	PPB	26 Mar 2021 11:47:05	S	Std	
2.5 - 1	52.5	9315	-	2.50007470A	PPB	26 Mar 2021 11:49:22	S	Std	
5.0 - 1	55	20016	-	5.00007470A	PPB	26 Mar 2021 11:51:39	S	Std	
7.5 - 1	57.5	30329	-	7.50007470A	PPB	26 Mar 2021 11:53:55	S	Std	
10.0 - 1	57.6	39591	-	10.00007470A	PPB	26 Mar 2021 11:56:10	S	Std	
ICV54 - 1	ICV54	14702	3.7084	-7470A	PPB	26 Mar 2021 11:59:02	U	SMPL	
ICB54 - 1	ICB54	-43	0.0243	-7470A	PPB	26 Mar 2021 12:01:16	U	SMPL	
CCV86 - 1	CCV86	20728	5.2141	-7470A	PPB	26 Mar 2021 12:06:01	U	SMPL	
CCB86 - 1	CCB86	-51	0.0223	-7470A	PPB	26 Mar 2021 12:08:15	U	SMPL	
CRA - 1	CRA	738	0.2194	-7470A	PPB	26 Mar 2021 12:10:30	U	SMPL	
HighStd - 1	HighStd	38568	9.6716	-7470A	PPB	26 Mar 2021 12:12:44	U	SMPL	
ChkStd - 1	ChkStd	26149	6.5686	-7470A	PPB	26 Mar 2021 12:14:59	U	SMPL	
PB135360BL - 1	PBW	-77	0.0158	-7470A	PPB	26 Mar 2021 12:17:15	U	SMPL	
PB135360BS - 1	LCSW	13801	3.4833	-7470A	PPB	26 Mar 2021 12:19:31	U	SMPL	
M1791-16 - 1	EXTERIOR-SLAB-BRICK-1	-17	0.0308	-7470A	PPB	26 Mar 2021 12:21:48	U	SMPL	
M1791-19 - 1	EXTERIOR-SLAB-BRICK-2	33	0.0433	-7470A	PPB	26 Mar 2021 12:24:03	U	SMPL	
M1791-22 - 1	GRAY-INTERIOR-BLOCK-1-2	38	0.0445	-7470A	PPB	26 Mar 2021 12:26:20	U	SMPL	
M1791-27 - 1	ROSE-BRICK-INTERIOR	21	0.0403	-7470A	PPB	26 Mar 2021 12:28:37	U	SMPL	
M1792-03 - 1	TP-5	37	0.0443	-7470A	PPB	26 Mar 2021 12:30:53	U	SMPL	
CCV87 - 1	CCV87	18867	4.7491	-7470A	PPB	26 Mar 2021 12:33:08	U	SMPL	
CCB87 - 1	CCB87	-39	0.0253	-7470A	PPB	26 Mar 2021 12:35:24	U	SMPL	
M1792-06 - 1	TP-6	27	0.0418	-7470A	PPB	26 Mar 2021 12:37:39	U	SMPL	
M1796-01 - 1	PB-2021-WS-000-06	25	0.0413	-7470A	PPB	26 Mar 2021 12:39:54	U	SMPL	
M1796-01DUP - 1	PB-2021-WS-000-06DUP	33	0.0433	-7470A	PPB	26 Mar 2021 12:42:09	U	SMPL	
M1796-01MS - 1	PB-2021-WS-000-06MS	14384	3.6290	-7470A	PPB	26 Mar 2021 12:44:24	U	SMPL	
M1796-01MSD - 1	PB-2021-WS-000-06MSD	14551	3.6707	-7470A	PPB	26 Mar 2021 12:46:40	U	SMPL	
PB135365BL - 1	PBW	-28	0.0280	-7470A	PPB	26 Mar 2021 12:48:56	U	SMPL	
PB135365BS - 1	LCSW	13190	3.3307	-7470A	PPB	26 Mar 2021 12:51:13	U	SMPL	
M1770-20 - 1	MW-103S	135	0.0688	-7470A	PPB	26 Mar 2021 12:53:29	U	SMPL	
M1770-20DUP - 1	MW-103SDUP	187	0.0818	-7470A	PPB	26 Mar 2021 12:55:46	U	SMPL	
M1770-20MS - 1	MW-103SMS	16203	4.0835	-7470A	PPB	26 Mar 2021 12:58:02	U	SMPL	
CCV88 - 1	CCV88	20504	5.1581	-7470A	PPB	26 Mar 2021 13:00:17	U	SMPL	
CCB88 - 1	CCB88	-88	0.0131	-7470A	PPB	26 Mar 2021 13:02:32	U	SMPL	
M1770-20MSD - 1	MW-103SMSD	16209	4.0850	-7470A	PPB	26 Mar 2021 13:04:48	U	SMPL	
M1770-21 - 1	MW-102D	1779	0.4795	-7470A	PPB	26 Mar 2021 13:07:03	U	SMPL	
M1770-22 - 1	MW-102S	-7	0.0333	-7470A	PPB	26 Mar 2021 13:09:18	U	SMPL	
M1770-23 - 1	MW-102I	112	0.0630	-7470A	PPB	26 Mar 2021 13:11:34	U	SMPL	
M1770-24 - 1	MW-107I	98	0.0595	-7470A	PPB	26 Mar 2021 13:13:49	U	SMPL	
M1809-05 - 1	DRUM-4	987	0.2816	-7470A	PPB	26 Mar 2021 13:16:05	U	SMPL	
PB135360TB - 1	PB135333TB	5	0.0363	-7470A	PPB	26 Mar 2021 13:18:21	U	SMPL	
CCV89 - 1	CCV89	19530	4.9148	-7470A	PPB	26 Mar 2021 13:20:38	U	SMPL	
CCB89 - 1	CCB89	-12	0.0320	-7470A	PPB	26 Mar 2021 13:22:55	U	SMPL	
M1796-01LX5 - 1		28	0.0420	-7470A	PPB	26 Mar 2021 13:25:11	U	SMPL	
M1796-01A - 1		14401	3.6332	-7470A	PPB	26 Mar 2021 13:27:26	U	SMPL	
M1770-20LX5 - 1		11	0.0378	-7470A	PPB	26 Mar 2021 13:29:41	U	SMPL	
M1770-20A - 1		15219	3.8376	-7470A	PPB	26 Mar 2021 13:31:56	U	SMPL	
CCV90 - 1	CCV90	19280	4.8523	-7470A	PPB	26 Mar 2021 13:34:12	U	SMPL	
CCB90 - 1	CCB90	-28	0.0280	-7470A	PPB	26 Mar 2021 13:36:27	U	SMPL	