



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113696

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	03/26/2021	09:25
2	M1776-03		1	10	100.00	>212.0	>212.0	03/26/2021	09:45
3	M1776-03DUP		1	10	100.00	>212.0	>212.0	03/26/2021	10:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 113696

WorkList Name : fp32621-1 WorkList ID : 147505 Department : Wet-Chemistry Date : 03-26-2021 08:54:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/30/2021	Solid	M1776-03	Flash Point ✓	Cool 4 deg C	PSEG03	J43	106	03/23/2021	1010A

Date/Time 03/26/21 09:05
 Received by: KS (Wetchem)
 Relinquished by: SP (COT-HCM)

Date/Time 03/26/21 10:15
 Received by: SP (COT-HCM)
 Relinquished by: KS (Wetchem)