



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113707

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	03/26/2021	14:30
2	CAL2	1	Water	NA	NA	21.3	7.01	03/26/2021	14:31
3	CAL3	1	Water	NA	NA	21.3	10.03	03/26/2021	14:32
4	ICV	1	Water	NA	NA	21.4	7.00	03/26/2021	14:33
5	CCV1	1	Water	NA	NA	21.3	2.01	03/26/2021	14:35
6	M1825-01	1	Water	NA	NA	21.1	6.62	03/26/2021	15:20
7	M1825-01DUP	1	Water	NA	NA	21.2	6.63	03/26/2021	15:21
8	CCV2	1	Water	NA	NA	21.4	12.02	03/26/2021	15:24

2B 113207

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PHM1825 W WorkList ID : 147522 Department : Wet-Chemistry Date : 03-26-2021 15:28:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/26/2021	Water	M1825-01	pH	Cool 4 deg C	PSEG03	L13	FM5210	03/26/2021	9040C

Date/Time 03/26/21 15:30
Received by: JG Wet Chem
Relinquished by: JG Wet Chem

Date/Time 03/26/21 16:30
Received by: JG Wet Chem
Relinquished by: JG Wet Chem