



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113715

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/29/2021	12:05
2	CAL2	1	Water	NA	NA	20.2	7.00	03/29/2021	12:06
3	CAL3	1	Water	NA	NA	20.2	10.03	03/29/2021	12:07
4	ICV	1	Water	NA	NA	20.3	6.99	03/29/2021	12:10
5	CCV1	1	Water	NA	NA	20.3	2.01	03/29/2021	12:11
6	M1861-01	1	Water	NA	NA	20.4	7.60	03/29/2021	12:15
7	M1861-01DUP	1	Water	NA	NA	20.3	7.61	03/29/2021	12:16
8	CCV2	1	Water	NA	NA	20.4	12.02	03/29/2021	12:18

203 113715

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m1862 WorkList ID : 147549 Department : Wet-Chemistry Date : 03-29-2021 11:57:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/26/2021	Water	M1861-01	pH	Cool 4 deg C	MINO01	L11	021-OUTFAL	03/26/2021	9040C

Date/Time 03/29/21 12:00
Received by: JO /vet (mm)
Relinquished by: JO csm

Date/Time 03/29/21 13:30
Received by: JO csm
Relinquished by: JO csm

mm