



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113716

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/29/2021	12:05
2	CAL2	1	Water	NA	NA	20.2	7.00	03/29/2021	12:06
3	CAL3	1	Water	NA	NA	20.2	10.03	03/29/2021	12:07
4	ICV	1	Water	NA	NA	20.3	6.99	03/29/2021	12:10
5	CCV1	1	Water	NA	NA	20.3	2.01	03/29/2021	12:11
6	M1862-01	1	Water	NA	NA	20.3	6.09	03/29/2021	12:20
7	M1862-01DUP	1	Water	NA	NA	20.4	6.10	03/29/2021	12:21
8	M1862-02	1	Water	NA	NA	20.3	6.72	03/29/2021	12:24
9	M1862-03	1	Water	NA	NA	20.5	6.72	03/29/2021	12:25
10	M1862-04	1	Water	NA	NA	20.5	6.73	03/29/2021	12:26
11	CCV2	1	Water	NA	NA	20.3	12.03	03/29/2021	12:30

AP 113716

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m1861 WorkList ID : 147550 Department : Wet-Chemistry Date : 03-29-2021 11:58:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/25/2021	Water	M1862-01	pH	Cool 4 deg C	LOCK01	L11	MW-1	03/25/2021	SM 4500-H B
03/25/2021	Water	M1862-02	pH	Cool 4 deg C	LOCK01	L11	MW-3	03/25/2021	SM 4500-H B
03/25/2021	Water	M1862-03	pH	Cool 4 deg C	LOCK01	L11	MW-4	03/25/2021	SM 4500-H B
03/25/2021	Water	M1862-04	pH	Cool 4 deg C	LOCK01	L11	MW-2	03/25/2021	SM 4500-H B

Date/Time 03/29/21 12:00
 Received by: JD (wet chem)
 Relinquished by: JD csmj

Date/Time 03/29/21 13:30
 Received by: JD csmj
 Relinquished by: JD (wet chem)