


7471B

LB113725
INSTRUMENT ID: CV1

Linear

 μ Abs.:

56886



A= 0.0000e+000

B= 1.7662e-004

C= 1.7283e-003

Rho= 0.9999559

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.013	0.013	63	0.000	63					
0.2	0.200	0.222	0.022	1245	0.0 %	1245					
2.5	2.500	2.474	-0.026	13998	0.0 %	13998					
5.0	5.000	5.002	0.002	28311	0.0 %	28311					
7.5	7.500	7.441	-0.059	42119	0.0 %	42119					
10.0	10.000	10.049	0.049	56886	0.0 %	56886					

900
11
-1
-0
-1
-0

LB113725 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 29 Mar 2021 15:51:23				
Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type
0.0 - 1	50	63	-	0.0000	7471B	PPB	29 Mar 2021 15:52:28	S Std
0.2 - 1	50.2	1245	-	0.2000	7471B	PPB	29 Mar 2021 15:54:44	S Std
2.5 - 1	52.5	13998	-	2.5000	7471B	PPB	29 Mar 2021 15:57:01	S Std
5.0 - 1	55	28311	-	5.0000	7471B	PPB	29 Mar 2021 15:59:18	S Std
7.5 - 1	57.5	42119	-	7.5000	7471B	PPB	29 Mar 2021 16:01:34	S Std
10.0 - 1	510	56886	-	10.0000	7471B	PPB	29 Mar 2021 16:03:50	S Std
ICV57 - 1	ICV57	21818	3.8552	-	7471B	PPB	29 Mar 2021 16:06:31	U SMPL
ICB57 - 1	ICB57	-4	0.0010	-	7471B	PPB	29 Mar 2021 16:08:45	U SMPL
CCV01 - 1	CCV01	29033	5.1295	-	7471B	PPB	29 Mar 2021 16:11:00	U SMPL
CCB01 - 1	CCB01	-23	-0.0023	-	7471B	PPB	29 Mar 2021 16:13:15	U SMPL
CRA - 1	CRA	1232	0.2193	-	7471B	PPB	29 Mar 2021 16:15:30	U SMPL
HighStd - 1	HighStd	56779	10.0299	-	7471B	PPB	29 Mar 2021 16:17:45	U SMPL
ChkStd - 1	ChkStd	39234	6.9312	-	7471B	PPB	29 Mar 2021 16:20:00	U SMPL
PB135408BL - 1	PBS	-71	-0.0108	-	7471B	PPB	29 Mar 2021 16:22:17	U SMPL
PB135408BS - 1	LCSS	20086	3.5493	-	7471B	PPB	29 Mar 2021 16:24:34	U SMPL
M1821-01 - 1	EL-509-STREET-LEVEL	262	0.0480	-	7471B	PPB	29 Mar 2021 16:26:50	U SMPL
M1821-01DUP - 1	EL-509-STREET-LEVELDUP	127	0.0242	-	7471B	PPB	29 Mar 2021 16:29:07	U SMPL
M1821-01MS - 1	EL-509-STREET-LEVELMS	20613	3.6424	-	7471B	PPB	29 Mar 2021 16:31:24	U SMPL
M1821-01MSD - 1	EL-509-STREET-LEVELMSD	20747	3.6660	-	7471B	PPB	29 Mar 2021 16:33:40	U SMPL
M1821-02 - 1	EL-510-STREET-LEVEL	5453	0.9648	-	7471B	PPB	29 Mar 2021 16:35:56	U SMPL
CCV02 - 1	CCV02	28724	5.0749	-	7471B	PPB	29 Mar 2021 16:38:11	U SMPL
CCB02 - 1	CCB02	-15	-0.0009	-	7471B	PPB	29 Mar 2021 16:40:26	U SMPL
M1821-03 - 1	EL-511-STREET-LEVEL	23833	4.2111	-	7471B	PPB	29 Mar 2021 16:42:42	U SMPL
M1826-01 - 1	RO-RBR200035	4978	0.8809	-	7471B	PPB	29 Mar 2021 16:44:58	U SMPL
M1828-01 - 1	FS-SB13(5-6)	4836	0.8558	-	7471B	PPB	29 Mar 2021 16:47:13	U SMPL
M1828-02 - 1	FS-SB13(25-26)	161	0.0302	-	7471B	PPB	29 Mar 2021 16:49:28	U SMPL
M1828-03 - 1	FS-SB13(33-34)	170	0.0318	-	7471B	PPB	29 Mar 2021 16:51:44	U SMPL
M1828-04 - 1	FS-SB14(5-6)	8634	1.5266	-	7471B	PPB	29 Mar 2021 16:54:00	U SMPL
M1828-05 - 1	FS-SB14(15-16)	3188	0.5648	-	7471B	PPB	29 Mar 2021 16:56:17	U SMPL
M1832-01 - 1	IG001	2929	0.5190	-	7471B	PPB	29 Mar 2021 16:58:33	U SMPL
M1832-02 - 1	IG002	2731	0.4841	-	7471B	PPB	29 Mar 2021 17:00:50	U SMPL
M1832-03 - 1	IG003	2057	0.3650	-	7471B	PPB	29 Mar 2021 17:03:06	U SMPL
CCV03 - 1	CCV03	29042	5.1311	-	7471B	PPB	29 Mar 2021 17:05:21	U SMPL
CCB03 - 1	CCB03	-8	0.0003	-	7471B	PPB	29 Mar 2021 17:07:37	U SMPL
M1832-04 - 1	IG004	2929	0.5190	-	7471B	PPB	29 Mar 2021 17:09:52	U SMPL
M1832-05 - 1	IG005	7282	1.2879	-	7471B	PPB	29 Mar 2021 17:12:07	U SMPL
M1838-01 - 1	PHC-BEAPT-001	13231	2.3386	-	7471B	PPB	29 Mar 2021 17:14:23	U SMPL
M1855-01 - 1	EMBANKMENT-FILL-COMPOSITE	3188	0.5648	-	7471B	PPB	29 Mar 2021 17:16:38	U SMPL
M1821-01LX5 - 1		34	0.0077	-	7471B	PPB	29 Mar 2021 17:18:54	U SMPL
M1821-01A - 1		21668	3.8287	-	7471B	PPB	29 Mar 2021 17:21:10	U SMPL
CCV04 - 1	CCV04	29136	5.1477	-	7471B	PPB	29 Mar 2021 17:23:26	U SMPL
CCB04 - 1	CCB04	-5	0.0008	-	7471B	PPB	29 Mar 2021 17:25:43	U SMPL