



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB113734

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP90232
Turbidity Calibration std, 20NTU	WP90228
Turbidity Calibration std, 40NTU	WP90227
Turbidity Calibration std, 1NTU	WP90231
Turbidity Calibration std, 5NTU	WP90229
Turbidity Calibration - CCV std, 10 NTU	WP90230
10 NTU Standard 500 ml	W2752

Intercept: 0.2266

Slope: 0.9856

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.234	0.01		03/29/2021	11:15
2	CAL2	1	1	1.170	0.96	4.3	03/29/2021	11:17
3	CAL3	5	1	5.240	5.09	-1.7	03/29/2021	11:19
4	CAL4	10	1	9.920	9.84	1.6	03/29/2021	11:21
5	CAL5	20	1	20.100	20.16	-0.8	03/29/2021	11:24
6	CAL6	40	1	39.600	39.95	0.1	03/29/2021	11:26



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB113734

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.200	10.119	03/29/2021	11:29
2	ICB		1	0.248	0.022	03/29/2021	11:31
3	CCV1	10	1	9.800	9.713	03/29/2021	11:33
4	CCB1		1	0.252	0.026	03/29/2021	11:35
5	LB113734BL		1	0.261	0.035	03/29/2021	11:38
6	M1862-01		1	0.580	0.359	03/29/2021	11:40
7	M1862-01DUP		1	0.610	0.389	03/29/2021	11:42
8	M1862-02		1	0.454	0.231	03/29/2021	11:44
9	M1862-03		1	0.383	0.159	03/29/2021	11:47
10	M1862-04		1	13.600	13.569	03/29/2021	11:50
11	CCV2	10	1	10.100	10.018	03/29/2021	11:52
12	CCB2		1	0.286	0.060	03/29/2021	11:54

WORKLIST(Hardcopy Internal Chain)

LB113733-LB113734

WorkList Name : Turbidity-32921 WorkList ID : 147555 Department : Wet-Chemistry Date : 03-29-2021 12:07:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/27/2021	Water	M1862-01 F	Turbidity	Cool 4 deg C	LOCK01	L11	MW-1	03/25/2021	SM2130 B
03/27/2021	Water	M1862-02 F	Turbidity	Cool 4 deg C	LOCK01	L11	MW-3	03/25/2021	SM2130 B
03/27/2021	Water	M1862-03 F	Turbidity	Cool 4 deg C	LOCK01	L11	MW-4	03/25/2021	SM2130 B
03/27/2021	Water	M1862-04 F	Turbidity	Cool 4 deg C	LOCK01	L11	MW-2	03/25/2021	SM2130 B

Date/Time 03/29/2021 10:55
 Received by: AM (wetchem)
 Relinquished by: JD (C694) (Hem)

Date/Time 03/29/2021 12:30
 Received by: JD (C694) (Hem)
 Relinquished by: AM (wetchem)