



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113746

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/30/2021	11:30
2	CAL2	1	Water	NA	NA	20.3	7.00	03/30/2021	11:31
3	CAL3	1	Water	NA	NA	20.3	10.03	03/30/2021	11:32
4	ICV	1	Water	NA	NA	20.4	6.99	03/30/2021	11:35
5	CCV1	1	Water	NA	NA	20.4	2.01	03/30/2021	11:36
6	M1869-02	1	Water	NA	NA	20.3	6.15	03/30/2021	11:39
7	M1869-02DUP	1	Water	NA	NA	20.4	6.16	03/30/2021	11:40
8	M1869-03	1	Water	NA	NA	20.4	7.31	03/30/2021	11:42
9	CCV2	1	Water	NA	NA	20.3	12.02	03/30/2021	11:43

LB 113446

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M1869

WorkList ID : 147595

Department : Wet-Chemistry

Date : 03-30-2021 11:21:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/29/2021	Water	M1869-02	pH	Cool 4 deg C	TETR06	L11	TT-029-IDWGW-20210329	03/29/2021	9040C
03/29/2021	Water	M1869-03	pH	Cool 4 deg C	TETR06	L11	TT-030-IDWGW-20210329	03/29/2021	9040C

Date/Time 03/30/21 11:25

Received by: JO (wet Chem)

Relinquished by: JO (Nettack)

Date/Time 03/30/21 12:30

Received by: JO (Nettack)

Relinquished by: JO (wet Chem)