

LB113749


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51852



A= 0.0000e+000

B= 1.9502e-004

C= -4.9418e-002

Rho= 0.9999260

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.029	0.029	404	0.000	404					
0.2	0.200	0.213	0.013	1348	0.0 %	1348					
2.5	2.500	2.468	-0.032	12910	0.0 %	12910					
5.0	5.000	5.004	0.004	25910	0.0 %	25910					
7.5	7.500	7.422	-0.078	38311	0.0 %	38311					
10.0	10.000	10.063	0.063	51852	0.0 %	51852					

LB113749 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 Mar 2021 13:27:08

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	404	-	0.0000	7471B	PPB	30 Mar 2021 13:28:58	S	Std
0.2 - 1	50.2	1348	-	0.2000	7471B	PPB	30 Mar 2021 13:31:15	S	Std
2.5 - 1	52.5	12910	-	2.5000	7471B	PPB	30 Mar 2021 13:33:32	S	Std
5.0 - 1	55	25910	-	5.0000	7471B	PPB	30 Mar 2021 13:35:48	S	Std
7.5 - 1	57.5	38311	-	7.5000	7471B	PPB	30 Mar 2021 13:40:43	S	Std
10.0 - 1	50	51852	-	10.0000	7471B	PPB	30 Mar 2021 13:42:59	S	Std
ICV59 - 1	ICV59	19421	3.7381	-	7471B	PPB	30 Mar 2021 13:45:44	U	SMPL
ICB59 - 1	ICB59	2	-0.0490	-	7471B	PPB	30 Mar 2021 13:47:59	U	SMPL
CCV08 - 1	CCV08	24712	4.7700	-	7471B	PPB	30 Mar 2021 13:50:16	U	SMPL
CCB08 - 1	CCB08	241	-0.0024	-	7471B	PPB	30 Mar 2021 13:55:02	U	SMPL
CRA - 1	CRA	1253	0.1949	-	7471B	PPB	30 Mar 2021 13:57:17	U	SMPL
HighStd - 1	HighStd	51750	10.0431	-	7471B	PPB	30 Mar 2021 13:59:32	U	SMPL
ChkStd - 1	ChkStd	36084	6.9878	-	7471B	PPB	30 Mar 2021 14:01:49	U	SMPL
PB135445BL - 1	PBS	23	-0.0449	-	7471B	PPB	30 Mar 2021 14:04:07	U	SMPL
PB135445BS - 1	LCSS	20938	4.0340	-	7471B	PPB	30 Mar 2021 14:06:24	U	SMPL
M1823-01 - 1	MK-3-29-B1	6686	1.2545	-	7471B	PPB	30 Mar 2021 14:08:40	U	SMPL
M1823-01DUP - 1	MK-3-29-B1DUP	6314	1.1820	-	7471B	PPB	30 Mar 2021 14:16:15	U	SMPL
M1823-01MS - 1	MK-3-29-B1MS	27068	5.2295	-	7471B	PPB	30 Mar 2021 14:18:31	U	SMPL
M1823-01MSD - 1	MK-3-29-B1MSD	27332	5.2810	-	7471B	PPB	30 Mar 2021 14:23:18	U	SMPL
M1836-01 - 1	B1-0-2	11354	2.1649	-	7471B	PPB	30 Mar 2021 14:25:33	U	SMPL
CCV09 - 1	CCV09	24497	4.7281	-	7471B	PPB	30 Mar 2021 14:27:49	U	SMPL
CCB09 - 1	CCB09	186	-0.0131	-	7471B	PPB	30 Mar 2021 14:30:04	U	SMPL
M1836-02 - 1	B1-10-12	1205	0.1856	-	7471B	PPB	30 Mar 2021 14:32:21	U	SMPL
M1836-03 - 1	B2-0-2	2334	0.4058	-	7471B	PPB	30 Mar 2021 14:34:36	U	SMPL
M1836-04 - 1	B2-10-12	2092	0.3586	-	7471B	PPB	30 Mar 2021 14:36:52	U	SMPL
M1836-05 - 1	B3-0-2	12503	2.3890	-	7471B	PPB	30 Mar 2021 14:39:08	U	SMPL
M1836-06 - 1	B3-10-12	69205	13.4472	-	7471B	PPB	30 Mar 2021 14:41:23	U	SMPL
M1836-07 - 1	B3-19-21	1088	0.1628	-	7471B	PPB	30 Mar 2021 14:43:39	U	SMPL
M1836-08 - 1	B4-0-2	7283	1.3709	-	7471B	PPB	30 Mar 2021 14:45:57	U	SMPL
M1836-09 - 1	B4-8-10	5446	1.0127	-	7471B	PPB	30 Mar 2021 14:48:14	U	SMPL
M1836-10 - 1	B5-0-2	12679	2.4233	-	7471B	PPB	30 Mar 2021 14:50:31	U	SMPL
M1836-11 - 1	B5-8-10	1402	0.2240	-	7471B	PPB	30 Mar 2021 14:52:46	U	SMPL
CCV10 - 1	CCV10	25264	4.8777	-	7471B	PPB	30 Mar 2021 14:55:02	U	SMPL
CCB10 - 1	CCB10	233	-0.0040	-	7471B	PPB	30 Mar 2021 14:57:17	U	SMPL
M1836-12 - 1	B5-19-21	2372	0.4132	-	7471B	PPB	30 Mar 2021 14:59:32	U	SMPL
M1836-13 - 1	B6-0-2	14943	2.8648	-	7471B	PPB	30 Mar 2021 15:01:48	U	SMPL
M1836-14 - 1	B6-10-12	1712	0.2845	-	7471B	PPB	30 Mar 2021 15:04:04	U	SMPL
M1836-15 - 1	B7-0-2	4071	0.7445	-	7471B	PPB	30 Mar 2021 15:06:19	U	SMPL
M1836-16 - 1	B7-10-12	769	0.1006	-	7471B	PPB	30 Mar 2021 15:08:35	U	SMPL
M1852-01 - 1	OR-03-032921	2908	0.5177	-	7471B	PPB	30 Mar 2021 15:10:52	U	SMPL
M1823-01LX5 - 1		1506	0.2443	-	7471B	PPB	30 Mar 2021 15:13:08	U	SMPL
M1823-01A - 1		28029	5.4169	-	7471B	PPB	30 Mar 2021 15:15:24	U	SMPL
M1836-06DLX2 - 1		32477	6.2844	-	7471B	PPB	30 Mar 2021 15:17:41	U	SMPL
CCV11 - 1	CCV11	24147	4.6598	-	7471B	PPB	30 Mar 2021 15:19:57	U	SMPL
CCB11 - 1	CCB11	214	-0.0077	-	7471B	PPB	30 Mar 2021 15:22:13	U	SMPL