



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB113751

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/30/2021	14:30
2	CAL2	1	Water	NA	NA	20.3	7.00	03/30/2021	14:31
3	CAL3	1	Water	NA	NA	20.3	10.03	03/30/2021	14:32
4	ICV	1	Water	NA	NA	20.4	6.99	03/30/2021	14:35
5	CCV1	1	Water	NA	NA	20.4	2.01	03/30/2021	14:36
6	M1864-01	1	Solid	20.03	20	20.6	9.78	03/30/2021	14:39
7	M1864-01DUP	1	Solid	20.04	20	20.7	9.79	03/30/2021	14:40
8	M1872-02	1	Solid	20.02	20	20.8	8.28	03/30/2021	14:42
9	CCV2	1	Water	NA	NA	20.3	12.02	03/30/2021	14:45

2013751

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY M1864 WorkList ID : 147591 Department : Wet-Chemistry Date : 03-30-2021 10:19:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/29/2021	Solid	M1864-01	Corrosivity	Cool 4 deg C	UNIO02	L11	MMR-SP	03/29/2021	9045D
03/30/2021	Solid	M1872-02	Corrosivity	Cool 4 deg C	PSEG03	L43	SP-GIS	03/30/2021	9045D

Date/Time 03/30/21 12:10
Received by: JO / 694 (mm)
Relinquished by: J.C. (mm)

Date/Time 03/30/21 16:10
Received by: J.C. (mm)
Relinquished by: JO / 694 (mm)