



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB113752

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/30/2021	14:30
2	CAL2	1	Water	NA	NA	20.3	7.00	03/30/2021	14:31
3	CAL3	1	Water	NA	NA	20.3	10.03	03/30/2021	14:32
4	ICV	1	Water	NA	NA	20.4	6.99	03/30/2021	14:35
5	CCV1	1	Water	NA	NA	20.4	2.01	03/30/2021	14:36
6	M1873-06	1	Water	NA	NA	20.5	6.77	03/30/2021	15:15
7	M1873-07	1	Water	NA	NA	20.4	6.51	03/30/2021	15:18
8	M1873-08	1	Water	NA	NA	20.6	9.50	03/30/2021	15:22
9	M1873-08DUP	1	Water	NA	NA	20.7	9.51	03/30/2021	15:23
10	CCV2	1	Water	NA	NA	20.4	12.02	03/30/2021	15:25

HP 113752

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m1873 WorkList ID : 147598 Department : Wet-Chemistry Date : 03-30-2021 15:03:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/30/2021	Water	M1873-06	✓ pH	Cool 4 deg C	PSEG05	L32	TOTE-A	03/30/2021	9040C
03/30/2021	Water	M1873-07	✓ pH	Cool 4 deg C	PSEG05	L32	TOTE-B	03/30/2021	9040C
03/30/2021	Water	M1873-08	✓ pH	Cool 4 deg C	PSEG05	L32	TOTE-C	03/30/2021	9040C

Date/Time 03/30/21 15:10
Received by: JP (Yot-Oil m)
Relinquished by: A. (w) (sur)

Date/Time 03/30/21 16:30
Received by: A.C. (sur)
Relinquished by: JP (w) Chen