

LB113753


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34522



A= 0.0000e+000

B= 2.9138e-004

C= -3.0725e-002

Rho= 0.9997797

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	V/D
0.0	0.000	0.009	0.009	136	0.000	136					2
0.2	0.200	0.203	0.003	802	0.0 %	802					3
2.5	2.500	2.585	0.085	8977	0.0 %	8977					3
5.0	5.000	4.837	-0.163	16707	0.0 %	16707					1
7.5	7.500	7.538	0.038	25975	0.0 %	25975					0
10.0	10.000	10.028	0.028	34522	0.0 %	34522					

LB113753 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Mar 2021 15:30:20

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	136	-	0.0000	7470A	PPB	30 Mar 2021 15:33:20	S	Std
0.2 - 1	51.2	802	-	0.2000	7470A	PPB	30 Mar 2021 15:35:36	S	Std
2.5 - 1	52.5	8977	-	2.5000	7470A	PPB	30 Mar 2021 15:37:53	S	Std
5.0 - 1	55	16707	-	5.0000	7470A	PPB	30 Mar 2021 15:40:10	S	Std
7.5 - 1	57.5	25975	-	7.5000	7470A	PPB	30 Mar 2021 15:42:26	S	Std
10.0 - 1	510	34522	-	10.0000	7470A	PPB	30 Mar 2021 15:47:15	S	Std
ICV60 - 1	ICV60	14980	4.3341	-	7470A	PPB	30 Mar 2021 15:50:29	U	SMPL
ICB60 - 1	ICB60	38	-0.0197	-	7470A	PPB	30 Mar 2021 15:55:13	U	SMPL
CCV12 - 1	CCV12	17243	4.9935	-	7470A	PPB	30 Mar 2021 16:00:02	U	SMPL
CCB12 - 1	CCB12	37	-0.0199	-	7470A	PPB	30 Mar 2021 16:02:17	U	SMPL
CRA - 1	CRA	845	0.2155	-	7470A	PPB	30 Mar 2021 16:04:32	U	SMPL
HighStd - 1	HighStd	35839	10.4119	-	7470A	PPB	30 Mar 2021 16:10:00	U	SMPL
ChkStd - 1	ChkStd	23440	6.7991	-	7470A	PPB	30 Mar 2021 16:14:49	U	SMPL
PB135452BL - 1	PBW	118	0.0037	-	7470A	PPB	30 Mar 2021 16:19:36	U	SMPL
PB135452BS - 1	LCSW	13606	3.9337	-	7470A	PPB	30 Mar 2021 16:21:52	U	SMPL
M1869-02 - 1	TT-029-IDWGW-20210329	11	-0.0275	-	7470A	PPB	30 Mar 2021 16:24:08	U	SMPL
M1869-02DUP - 1	TT-029-IDWGW-20210329DUP	161	0.0162	-	7470A	PPB	30 Mar 2021 16:26:25	U	SMPL
M1869-02MS - 1	TT-029-IDWGW-20210329MS	15410	4.4594	-	7470A	PPB	30 Mar 2021 16:28:42	U	SMPL
M1869-02MSD - 1	TT-029-IDWGW-20210329MSD	15686	4.5398	-	7470A	PPB	30 Mar 2021 16:30:59	U	SMPL
M1869-03 - 1	TT-030-IDWGW-20210329	3001	0.8437	-	7470A	PPB	30 Mar 2021 16:33:14	U	SMPL
CCV13 - 1	CCV13	17159	4.9690	-	7470A	PPB	30 Mar 2021 16:35:29	U	SMPL
CCB13 - 1	CCB13	78	-0.0080	-	7470A	PPB	30 Mar 2021 16:37:44	U	SMPL
M1869-02LX5 - 1		198	0.0270	-	7470A	PPB	30 Mar 2021 16:39:59	U	SMPL
M1869-02A - 1		14859	4.2988	-	7470A	PPB	30 Mar 2021 16:42:14	U	SMPL
CCV14 - 1	CCV14	18512	5.3632	-	7470A	PPB	30 Mar 2021 16:44:30	U	SMPL
CCB14 - 1	CCB14	-55	-0.0468	-	7470A	PPB	30 Mar 2021 16:46:45	U	SMPL