


LB 113759

ISM02.4

INSTRUMENT ID: CV2

 μ Abs.:

43480



Linear

A= 0.0000e+000

B= 2.2834e-004 *slope*C= 6.4172e-003 *y-intercept*

Rho= 0.9998737

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.029	0.029	99	0.000	99					=
0.05	0.050					204					-4
0.20	0.200	0.192	-0.008	811	0.0 %	811					-3
2.50	2.500	2.418	-0.082	10560	0.0 %	10560					1
5.00	5.000	5.045	0.045	22066	0.0 %	22066					1
7.50	7.500	7.582	0.082	33175	0.0 %	33175					-1
10.0	10.000	9.935	-0.065	43480	0.0 %	43480					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	99	-		0 ISM02.4	PPB	3/30/2021 16:55	Std
	0.2 S0.2	811	-		0.2 ISM02.4	PPB	3/30/2021 16:57	Std
	2.5 S2.5	10560	-		2.5 ISM02.4	PPB	3/30/2021 17:00	Std
	5 S5.0	22066	-		5 ISM02.4	PPB	3/30/2021 17:02	Std
	7.5 S7.5	33175	-		7.5 ISM02.4	PPB	3/30/2021 17:04	Std
	10 S10.0	43480	-		10 ISM02.4	PPB	3/30/2021 17:06	Std
ICV	ICV	16298	3.728 -		ISM02.4	PPB	3/30/2021 17:09	SMPL
ICB	ICB	-162	-0.0306 -		ISM02.4	PPB	3/30/2021 17:11	SMPL
CCV015	CCV015	22143	5.0626 -		ISM02.4	PPB	3/30/2021 17:14	SMPL
CCB015	CCB015	-130	-0.0233 -		ISM02.4	PPB	3/30/2021 17:16	SMPL
PB135440BL	PBW440	-9	0.0044 -		ISM02.4	PPB	3/30/2021 17:18	SMPL
M1819-01	MCOAA0	16	0.0101 -		ISM02.4	PPB	3/30/2021 17:20	SMPL
M1819-02	MCOAA1	17	0.0103 -		ISM02.4	PPB	3/30/2021 17:23	SMPL
M1819-03	MCOAA2	13	0.0094 -		ISM02.4	PPB	3/30/2021 17:25	SMPL
M1819-04	MCOAA3	6	0.0078 -		ISM02.4	PPB	3/30/2021 17:27	SMPL
M1819-05	MCOAA3D	8	0.0082 -		ISM02.4	PPB	3/30/2021 17:29	SMPL
M1819-06	MCOAA3S	3989	0.9173 -		ISM02.4	PPB	3/30/2021 17:32	SMPL
CCV016	CCV016	21848	4.9953 -		ISM02.4	PPB	3/30/2021 17:34	SMPL
CCB016	CCB016	-101	-0.0166 -		ISM02.4	PPB	3/30/2021 17:36	SMPL