



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB113772

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	03/31/2021	16:00
2	CAL2	1	Water	NA	NA	21.2	7.00	03/31/2021	16:01
3	CAL3	1	Water	NA	NA	21.2	10.03	03/31/2021	16:02
4	ICV	1	Water	NA	NA	21.2	6.99	03/31/2021	16:05
5	CCV1	1	Water	NA	NA	21.2	2.01	03/31/2021	16:06
6	M1883-02	1	Solid	20.03	20	21.1	6.74	03/31/2021	16:10
7	M1883-04	1	Solid	20.04	20	21.2	6.64	03/31/2021	16:11
8	M1883-04DUP	1	Solid	20.02	20	21.2	6.65	03/31/2021	16:15
9	CCV2	1	Water	NA	NA	21.3	12.02	03/31/2021	16:19

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY M1883 WorkList ID : 147631 Department : Wet-Chemistry Date : 03-31-2021 15:02:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/31/2021	Solid	M1883-02	Corrosivity	Cool 4 deg C	PSEG03	L21	IEC-R2	03/31/2021	9045D
03/31/2021	Solid	M1883-04	Corrosivity	Cool 4 deg C	PSEG03	L21	V-211	03/31/2021	9045D

Date/Time 03/31/21 15:10
 Received by: JR (chem)
 Relinquished by: JR (sm)

Date/Time 03-31-21 17:00
 Received by: JR (sm)
 Relinquished by: JR (sm)