



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB113778

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	04/01/2021	11:40
2	M1873-09		1	10	100.00	>212.0	>212.0	04/01/2021	12:05
3	M1873-09DUP		1	10	100.00	>212.0	>212.0	04/01/2021	12:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB113778

WorkList Name : fp4121

WorkList ID : 147649

Department : Wet-Chemistry

Date : 04-01-2021 09:00:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/02/2021	Water	M1873-09 C	Flash Point	Cool 4 deg C	PSEG03	L32	TOTE-COMP	03/30/2021	1010A

Date/Time 04/01/21 11:15
 Received by: KS (Wetchem)
 Relinquished by: AM (Wet-Chem)

Date/Time 04/01/21 12:40
 Received by: AM (Wetchem)
 Relinquished by: KS (Wetchem)