


L3113779

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.7291e-004

C= 3.0325e-002

Rho= 0.9999480

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.038	0.038	44	0.000	44					
0.2	0.200	0.209	0.009	1036	0.0 %	1036					
2.5	2.500	2.422	-0.078	13829	0.0 %	13829					
5.0	5.000	5.027	0.027	28897	0.0 %	28897					
7.5	7.500	7.496	-0.004	43178	0.0 %	43178					
10.0	10.000	10.008	0.008	57704	0.0 %	57704					

5
-3
1
0
0

LB113779 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 01 Apr 2021 11:15:26				
Sample ID	Extended ID	µ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	44	-	0.00007471B	PPB	01 Apr 2021 11:19:19	S	Std
0.2 - 1	50.2	1036	-	0.20007471B	PPB	01 Apr 2021 11:21:36	S	Std
2.5 - 1	52.5	13829	-	2.50007471B	PPB	01 Apr 2021 11:23:53	S	Std
5.0 - 1	55	28897	-	5.00007471B	PPB	01 Apr 2021 11:26:09	S	Std
7.5 - 1	57.5	43178	-	7.50007471B	PPB	01 Apr 2021 11:28:26	S	Std
10.0 - 1	510	57704	-	10.00007471B	PPB	01 Apr 2021 11:33:16	S	Std
ICV64 - 1	ICV64	22529	3.9258	-7471B	PPB	01 Apr 2021 11:35:57	U	SMPL
ICB64 - 1	ICB64	-203	-0.0048	-7471B	PPB	01 Apr 2021 11:38:11	U	SMPL
CCV21 - 1	CCV21	30003	5.2182	-7471B	PPB	01 Apr 2021 11:40:28	U	SMPL
CCB21 - 1	CCB21	-217	-0.0072	-7471B	PPB	01 Apr 2021 11:42:43	U	SMPL
CRA - 1	CRA	1037	0.2096	-7471B	PPB	01 Apr 2021 11:44:59	U	SMPL
HighStd - 1	HighStd	59076	10.2452	-7471B	PPB	01 Apr 2021 11:47:14	U	SMPL
ChkStd - 1	ChkStd	42550	7.3877	-7471B	PPB	01 Apr 2021 11:49:29	U	SMPL
PB135469BL - 1	PBS	-314	-0.0240	-7471B	PPB	01 Apr 2021 11:51:47	U	SMPL
PB135469BS - 1	LCSS	20438	3.5643	-7471B	PPB	01 Apr 2021 11:54:04	U	SMPL
M1872-02 - 1	SP-GIS	16292	2.8474	-7471B	PPB	01 Apr 2021 11:56:20	U	SMPL
M1874-01 - 1	FS-SB18(9-10)	781	0.1654	-7471B	PPB	01 Apr 2021 12:03:26	U	SMPL
M1874-01DUP - 1	FS-SB18(9-10)DUP	616	0.1368	-7471B	PPB	01 Apr 2021 12:05:44	U	SMPL
M1874-01MS - 1	FS-SB18(9-10)MS	22580	3.9346	-7471B	PPB	01 Apr 2021 12:08:02	U	SMPL
M1874-01MSD - 1	FS-SB18(9-10)MSD	23966	4.1743	-7471B	PPB	01 Apr 2021 12:10:17	U	SMPL
CCV22 - 1	CCV22	30374	5.2823	-7471B	PPB	01 Apr 2021 12:12:32	U	SMPL
CCB22 - 1	CCB22	-116	0.0103	-7471B	PPB	01 Apr 2021 12:14:48	U	SMPL
M1874-02 - 1	FS-SB18(24-25)	663	0.1450	-7471B	PPB	01 Apr 2021 12:17:03	U	SMPL
M1874-03 - 1	FS-SB18(39-40)	238	0.0715	-7471B	PPB	01 Apr 2021 12:19:19	U	SMPL
M1874-04 - 1	FS-SB16(7-8)	613	0.1363	-7471B	PPB	01 Apr 2021 12:21:34	U	SMPL
M1874-05 - 1	FS-SB16(26-27)	454	0.1088	-7471B	PPB	01 Apr 2021 12:23:49	U	SMPL
M1874-06 - 1	FS-SB16(45-46)	578	0.1303	-7471B	PPB	01 Apr 2021 12:26:05	U	SMPL
M1874-07 - 1	FS-SB11(14-15)	254	0.0742	-7471B	PPB	01 Apr 2021 12:28:21	U	SMPL
M1874-08 - 1	FS-SB11(52-53)	350	0.0908	-7471B	PPB	01 Apr 2021 12:30:37	U	SMPL
M1874-01LX5 - 1		78	0.0438	-7471B	PPB	01 Apr 2021 12:32:54	U	SMPL
M1874-01A - 1		23107	4.0258	-7471B	PPB	01 Apr 2021 12:35:11	U	SMPL
CCV23 - 1	CCV23	30394	5.2858	-7471B	PPB	01 Apr 2021 12:37:27	U	SMPL
CCB23 - 1	CCB23	-65	0.0191	-7471B	PPB	01 Apr 2021 12:39:43	U	SMPL