



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB113791

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/02/2021	08:30
2	CAL2	1	Water	NA	NA	20.1	7.00	04/02/2021	08:31
3	CAL3	1	Water	NA	NA	20.1	10.03	04/02/2021	08:32
4	ICV	1	Water	NA	NA	20.2	6.99	04/02/2021	08:35
5	CCV1	1	Water	NA	NA	20.1	2.01	04/02/2021	08:36
6	M1897-02	1	Solid	20.03	20	21.2	6.06	04/02/2021	08:38
7	M1897-02DUP	1	Solid	20.04	20	21.2	6.07	04/02/2021	08:39
8	M1897-04	1	Solid	20.02	20	21.2	6.26	04/02/2021	08:40
9	CCV2	1	Water	NA	NA	20.2	12.02	04/02/2021	08:42

WORKLIST(Hardcopy Internal Chain)

113791

WorkList Name : CORROSIONITY M1987 WorkList ID : 147676 Department : Wet-Chemistry Date : 04-02-2021 08:21:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/01/2021	Solid	M1897-02	Corrosivity	Cool 4 deg C	PSEG03	O11	TP-1	04/01/2021	9045D
04/01/2021	Solid	M1897-04	Corrosivity	Cool 4 deg C	PSEG03	O11	TP-2	04/01/2021	9045D

Date/Time 04/02/21 08:25
 Received by: SO (Wet-Chem)
 Relinquished by: KS (Wet-Chem)

Date/Time 04/02/21 11:00
 Received by: KS (Wet-Chem)
 Relinquished by: SO (Wet-Chem)